

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022522\
 Data File : VN071075.D
 Acq On : 25 Feb 2022 17:22
 Operator : JC\MD
 Sample : N1664-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3

Quant Time: Feb 26 03:43:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	852381	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1351825	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1211613	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	415556	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	511019	52.483	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.960%
35) Dibromofluoromethane	8.022	113	389022	48.382	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.760%
50) Toluene-d8	10.439	98	1664737	51.555	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.100%
62) 4-Bromofluorobenzene	12.727	95	561757	48.670	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.340%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	55273	10.308	ug/l	92
20) Methylene Chloride	5.098	84	111983	11.491	ug/l	91
43) Isopropyl Acetate	8.557	43	136525	6.131	ug/l #	79
95) Naphthalene	15.498	128	6914	5.325	ug/l	96
96) 1,2,3-Trichlorobenzene	15.704	180	2248	2.819	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022522\
Data File : VN071075.D
Acq On : 25 Feb 2022 17:22
Operator : JC\MD
Sample : N1664-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3

Quant Time: Feb 26 03:43:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

