

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022522\
 Data File : VN071079.D
 Acq On : 25 Feb 2022 18:58
 Operator : JC\MD
 Sample : N1664-20
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-7

Quant Time: Feb 26 03:44:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Feb 21 07:57:15 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	846924	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1379909	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1234377	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	424615	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	516472	53.385	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.760%
35) Dibromofluoromethane	8.022	113	396771	48.341	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.680%
50) Toluene-d8	10.439	98	1681796	51.023	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.040%
62) 4-Bromofluorobenzene	12.727	95	567309	48.151	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.300%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	51359	9.639	ug/l #	84
18) Methyl Acetate	4.869	43	24020	1.474	ug/l	99
20) Methylene Chloride	5.092	84	108053	11.159	ug/l	91
43) Isopropyl Acetate	8.563	43	174748	7.687	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022522\
Data File : VN071079.D
Acq On : 25 Feb 2022 18:58
Operator : JC\MD
Sample : N1664-20
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-7

Quant Time: Feb 26 03:44:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021522W.M
Quant Title : SW846 8260
QLast Update : Mon Feb 21 07:57:15 2022
Response via : Initial Calibration

