

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022525\
 Data File : VN085852.D
 Acq On : 25 Feb 2025 13:12
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Feb 25 23:32:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	202503	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	381753	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	317010	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	112020	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	159632	60.803	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	121.600%
35) Dibromofluoromethane	8.165	113	141415	56.611	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	113.220%
50) Toluene-d8	10.565	98	450557	49.574	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.140%
62) 4-Bromofluorobenzene	12.847	95	121584	40.603	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	81.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022525\
Data File : VN085852.D
Acq On : 25 Feb 2025 13:12
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Feb 25 23:32:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

