

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022525\
 Data File : VN085853.D
 Acq On : 25 Feb 2025 13:36
 Operator : JC\MD
 Sample : Q1415-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB02222025

Quant Time: Feb 25 23:33:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	210441	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	393207	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	337792	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	121445	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	162083	59.408	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	118.820%
35) Dibromofluoromethane	8.165	113	142826	55.510	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	111.020%
50) Toluene-d8	10.565	98	459744	49.112	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.220%
62) 4-Bromofluorobenzene	12.847	95	124996	40.526	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	81.060%
Target Compounds						
16) Acetone	4.436	43	10019	11.606	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022525\
Data File : VN085853.D
Acq On : 25 Feb 2025 13:36
Operator : JC\MD
Sample : Q1415-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB02222025

Quant Time: Feb 25 23:33:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

