

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN022619\
 Data File : VN053953.D
 Acq On : 26 Feb 2019 14:24
 Operator : JC/SP
 Sample : K1597-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-26

Quant Time: Feb 27 02:58:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 05:37:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	573553	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1092531	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1059093	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	388949	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	384991	57.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.52%	
35) Dibromofluoromethane	7.59	113	355083	50.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.52%	
50) Toluene-d8	10.09	98	1376580	52.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.68%	
62) 4-Bromofluorobenzene	12.40	95	451155	51.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.98%	
Target Compounds						
20) Methylene Chloride	4.55	84	92422	13.889	ug/l	92
43) Isopropyl Acetate	8.17	43	38647	3.155	ug/l #	89
52) Toluene	10.15	92	31651	1.722	ug/l	99
68) m/p-Xylenes	11.62	106	48270	3.413	ug/l	95
69) o-Xylene	11.95	106	20679	1.488	ug/l	98
80) 1,3,5-Trimethylbenzene	12.73	105	40971	1.653	ug/l	96
84) 1,2,4-Trimethylbenzene	13.04	105	114943	4.645	ug/l	97
95) Naphthalene	15.13	128	14128	4.070	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN022619\
Data File : VN053953.D
Acq On : 26 Feb 2019 14:24
Operator : JC/SP
Sample : K1597-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-26

Quant Time: Feb 27 02:58:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 05:37:40 2019
Response via : Initial Calibration

