

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN022619\
 Data File : VN053959.D
 Acq On : 26 Feb 2019 17:04
 Operator : JC/SP
 Sample : PB117375ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB117375ZHE#02

Quant Time: Feb 27 03:10:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 05:37:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	606646	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1094935	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1066302	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	385923	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	381357	53.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.26%	
35) Dibromofluoromethane	7.59	113	354823	50.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.22%	
50) Toluene-d8	10.09	98	1388501	52.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.36%	
62) 4-Bromofluorobenzene	12.40	95	456637	52.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.00%	
Target Compounds						
20) Methylene Chloride	4.55	84	17103	2.430	ug/l	92
43) Isopropyl Acetate	8.17	43	34084	2.776	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN022619\
Data File : VN053959.D
Acq On : 26 Feb 2019 17:04
Operator : JC/SP
Sample : PB117375ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB117375ZHE#02

Quant Time: Feb 27 03:10:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 05:37:40 2019
Response via : Initial Calibration

