

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN022619\
 Data File : VN053961.D
 Acq On : 26 Feb 2019 17:58
 Operator : JC/SP
 Sample : PB117375ZHE#04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB117375ZHE#04

Quant Time: Feb 27 03:13:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 05:37:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	637815	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1160889	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1128477	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	405351	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	405965	54.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.60%	
35) Dibromofluoromethane	7.59	113	373012	49.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.38%	
50) Toluene-d8	10.09	98	1474250	52.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.52%	
62) 4-Bromofluorobenzene	12.40	95	481263	51.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.38%	
Target Compounds						
20) Methylene Chloride	4.56	84	17894	2.418	ug/l #	91
43) Isopropyl Acetate	8.17	43	41150	3.162	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN022619\
Data File : VN053961.D
Acq On : 26 Feb 2019 17:58
Operator : JC/SP
Sample : PB117375ZHE#04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB117375ZHE#04

Quant Time: Feb 27 03:13:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 05:37:40 2019
Response via : Initial Calibration

