

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN022619\
 Data File : VN053968.D
 Acq On : 26 Feb 2019 21:05
 Operator : JC/SP
 Sample : K1610-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WV-A

Quant Time: Feb 27 03:36:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 05:37:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	610390	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1111108	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1082188	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	415057	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	391834	54.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.52%	
35) Dibromofluoromethane	7.59	113	355121	49.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.84%	
50) Toluene-d8	10.09	98	1399752	52.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.66%	
62) 4-Bromofluorobenzene	12.40	95	473427	53.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.26%	
Target Compounds						
20) Methylene Chloride	4.55	84	119771	16.913	ug/l	92
43) Isopropyl Acetate	8.17	43	34091	2.737	ug/l #	91
95) Naphthalene	15.13	128	7622	3.646	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN022619\
Data File : VN053968.D
Acq On : 26 Feb 2019 21:05
Operator : JC/SP
Sample : K1610-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WV-A

Quant Time: Feb 27 03:36:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 05:37:40 2019
Response via : Initial Calibration

