

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022620\
 Data File : VN060250.D
 Acq On : 26 Feb 2020 14:49
 Operator : JC/MD
 Sample : PB127076TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127076TB

Quant Time: Feb 27 05:21:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	220716	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	349880	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	327830	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	158145	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	136763	50.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.84%	
35) Dibromofluoromethane	7.57	113	97848	45.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.00%	
50) Toluene-d8	10.08	98	431203	50.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.82%	
62) 4-Bromofluorobenzene	12.40	95	162041	52.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.14%	
Target Compounds						
16) Acetone	3.79	43	6439	5.496	ug/l	92
20) Methylene Chloride	4.52	84	8117	3.473	ug/l	95
43) Isopropyl Acetate	8.14	43	63846	12.752	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022620\
Data File : VN060250.D
Acq On : 26 Feb 2020 14:49
Operator : JC/MD
Sample : PB127076TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127076TB

Quant Time: Feb 27 05:21:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

