

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022620\
 Data File : VN060254.D
 Acq On : 26 Feb 2020 16:20
 Operator : JC/MD
 Sample : L1676-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RO-1-2

Quant Time: Feb 27 05:30:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 12 14:12:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	214645	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	342536	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	325602	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	160170	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	134985	51.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.36%	
35) Dibromofluoromethane	7.57	113	98711	46.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.76%	
50) Toluene-d8	10.08	98	422859	51.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.00%	
62) 4-Bromofluorobenzene	12.40	95	162426	53.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.62%	
Target Compounds						
16) Acetone	3.79	43	6878	6.037	ug/l	94
20) Methylene Chloride	4.53	84	9495	4.177	ug/l	95
43) Isopropyl Acetate	8.14	43	57981	11.829	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022620\
Data File : VN060254.D
Acq On : 26 Feb 2020 16:20
Operator : JC/MD
Sample : L1676-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RO-1-2

Quant Time: Feb 27 05:30:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N021220W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 12 14:12:21 2020
Response via : Initial Calibration

