

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022621\
 Data File : VN066043.D
 Acq On : 26 Feb 2021 11:45
 Operator : JC/MD
 Sample : VN0226MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0226MBL01

Quant Time: Feb 26 23:30:28 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	105182	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.547	114	171068	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	151899	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	59144	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	71459	48.58	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.16%
35) Dibromofluoromethane	7.550	113	51187	48.35	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.70%
50) Toluene-d8	10.055	98	208542	47.99	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.98%
62) 4-Bromofluorobenzene	12.373	95	73134	44.83	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.66%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022621\
Data File : VN066043.D
Acq On : 26 Feb 2021 11:45
Operator : JC/MD
Sample : VN0226MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0226MBL01

Quant Time: Feb 26 23:30:28 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

