

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022621\
 Data File : VN066048.D
 Acq On : 26 Feb 2021 14:02
 Operator : JC/MD
 Sample : M1504-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GW805-10-15-022521

Quant Time: Feb 26 23:32:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	115361	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	187843	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	170785	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	70086	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	76418	47.37	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.74%
35) Dibromofluoromethane	7.550	113	56657	48.73	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.46%
50) Toluene-d8	10.058	98	226653	47.50	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	95.00%
62) 4-Bromofluorobenzene	12.373	95	82102	45.83	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.66%
Target Compounds						
					Qvalue	
16) Acetone	3.791	43	3910	6.15	ug/l	97
30) Chloroform	7.335	83	9775	3.87	ug/l	97
47) Bromodichloromethane	9.370	83	4066	2.25	ug/l #	92
60) Dibromochloromethane	10.875	129	1069	0.80	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022621\
Data File : VN066048.D
Acq On : 26 Feb 2021 14:02
Operator : JC/MD
Sample : M1504-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GW805-10-15-022521

Quant Time: Feb 26 23:32:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

