

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022621\
 Data File : VN066052.D
 Acq On : 26 Feb 2021 15:34
 Operator : JC/MD
 Sample : PB134838TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134838TB

Manual Integrations
 APPROVED

MMDadoda
 3/1/2021 10:06:34 AM

Quant Time: Feb 26 23:34:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	118407	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	194403	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	176755	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	74709	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	80194	48.43	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.86%
35) Dibromofluoromethane	7.553	113	59558	49.50	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.00%
50) Toluene-d8	10.058	98	233726	47.32	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.64%
62) 4-Bromofluorobenzene	12.373	95	86299	46.55	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.10%
Target Compounds						
16) Acetone	3.798	43	4944	7.95	ug/l	Qvalue 97
20) Methylene Chloride	4.525	84	5677m	3.70	ug/l	
43) Isopropyl Acetate	8.126	43	7919	2.87	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022621\
Data File : VN066052.D
Acq On : 26 Feb 2021 15:34
Operator : JC/MD
Sample : PB134838TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134838TB

Manual Integrations
APPROVED

MMDadoda
3/1/2021 10:06:34 AM

Quant Time: Feb 26 23:34:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

