

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022621\
 Data File : VN066054.D
 Acq On : 26 Feb 2021 16:19
 Operator : JC/MD
 Sample : PB134837ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB134837ZHE#02

Quant Time: Feb 26 23:34:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	115464	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	187132	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	173839	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	72430	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	81106	50.23	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.46%	
35) Dibromofluoromethane	7.553	113	59268	51.17	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 102.34%	
50) Toluene-d8	10.058	98	230123	48.41	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 96.82%	
62) 4-Bromofluorobenzene	12.373	95	85759	48.05	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 96.10%	
Target Compounds						
						Qvalue
16) Acetone	3.801	43	5113	8.53	ug/l	98
18) Methyl Acetate	4.299	43	1872	1.40	ug/l #	65
20) Methylene Chloride	4.528	84	5262	3.51	ug/l #	87
43) Isopropyl Acetate	8.129	43	7809	2.94	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022621\
Data File : VN066054.D
Acq On : 26 Feb 2021 16:19
Operator : JC/MD
Sample : PB134837ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB134837ZHE#02

Quant Time: Feb 26 23:34:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

