

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022621\
 Data File : VN066063.D
 Acq On : 26 Feb 2021 19:52
 Operator : JC/MD
 Sample : M1491-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SA-01-022521

Quant Time: Feb 26 23:37:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 17 15:48:13 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	115932	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	186115	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	170554	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	72298	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.987	65	79499	49.04	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 98.08%			
35) Dibromofluoromethane	7.550	113	58143	50.48	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 100.96%			
50) Toluene-d8	10.058	98	226233	47.85	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 95.70%			
62) 4-Bromofluorobenzene	12.373	95	84632	47.68	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 95.36%			
Target Compounds					Qvalue	
16) Acetone	3.798	43	6414	11.06	ug/l	95
20) Methylene Chloride	4.528	84	4971	3.31	ug/l	87
43) Isopropyl Acetate	8.129	43	7443	2.82	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022621\
Data File : VN066063.D
Acq On : 26 Feb 2021 19:52
Operator : JC/MD
Sample : M1491-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SA-01-022521

Quant Time: Feb 26 23:37:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021721W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 17 15:48:13 2021
Response via : Initial Calibration

