

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022624\
 Data File : VN081189.D
 Acq On : 26 Feb 2024 12:52
 Operator : JC\MD
 Sample : P1533-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW8-MW01S-20240217

Quant Time: Feb 27 00:06:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	383336	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	701395	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	592158	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	227321	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	240168	51.115	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.220%
35) Dibromofluoromethane	8.165	113	193736	48.847	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.700%
50) Toluene-d8	10.565	98	751594	53.355	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.700%
62) 4-Bromofluorobenzene	12.847	95	236610	44.701	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022624\
Data File : VN081189.D
Acq On : 26 Feb 2024 12:52
Operator : JC\MD
Sample : P1533-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW8-MW01S-20240217

Quant Time: Feb 27 00:06:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

