

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022624\
 Data File : VN081190.D
 Acq On : 26 Feb 2024 13:16
 Operator : JC\MD
 Sample : P1533-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB04-20240217

Quant Time: Feb 27 00:06:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	408893	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	722625	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	619438	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	239651	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	248899	49.662	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.320%
35) Dibromofluoromethane	8.165	113	198225	48.511	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.020%
50) Toluene-d8	10.571	98	767072	52.854	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.700%
62) 4-Bromofluorobenzene	12.847	95	246421	45.187	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022624\
Data File : VN081190.D
Acq On : 26 Feb 2024 13:16
Operator : JC\MD
Sample : P1533-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB04-20240217

Quant Time: Feb 27 00:06:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

