

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022624\
 Data File : VN081191.D
 Acq On : 26 Feb 2024 13:40
 Operator : JC\MD
 Sample : P1576-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-195-EB-20240222

Quant Time: Feb 26 13:57:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	408448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	739342	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	637169	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	260171	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	264618	52.856	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	105.720%	
35) Dibromofluoromethane	8.171	113	208115	49.779	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.560%	
50) Toluene-d8	10.570	98	810014	54.550	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	109.100%	
62) 4-Bromofluorobenzene	12.853	95	266479	47.760	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	95.520%	
Target Compounds						
16) Acetone	4.436	43	5704	2.896	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022624\
Data File : VN081191.D
Acq On : 26 Feb 2024 13:40
Operator : JC\MD
Sample : P1576-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-195-EB-20240222

Quant Time: Feb 26 13:57:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

