

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022624\
 Data File : VN081195.D
 Acq On : 26 Feb 2024 15:16
 Operator : JC\MD
 Sample : P1533-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW9-MW01S-20240217

Quant Time: Feb 27 00:07:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	366542	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	665711	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	585094	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	228114	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	240052	53.431	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 106.860%			
35) Dibromofluoromethane	8.171	113	190702	50.659	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.320%			
50) Toluene-d8	10.571	98	737916	55.191	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 110.380%			
62) 4-Bromofluorobenzene	12.853	95	239202	47.613	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 95.220%			
Target Compounds						
29) Tetrahydrofuran	7.841	42	19955	10.467	ug/l	88
52) Toluene	10.635	92	21230	1.796	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022624\
Data File : VN081195.D
Acq On : 26 Feb 2024 15:16
Operator : JC\MD
Sample : P1533-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW9-MW01S-20240217

Quant Time: Feb 27 00:07:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

