

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022624\
 Data File : VN081202.D
 Acq On : 26 Feb 2024 18:04
 Operator : JC\MD
 Sample : P1577-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW10A-MW01S-20240221

Manual Integrations
 APPROVED

Reviewed By : John Carlone 02/27/2024
 Supervised By : Mahesh Dadoda 02/27/2024

Quant Time: Feb 27 00:14:34 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	346586	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	638618	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	564998	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	231082	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	229761	54.085	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery	= 108.160%		
35) Dibromofluoromethane	8.171	113	182813	50.624	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 101.240%		
50) Toluene-d8	10.571	98	714585	55.714	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery	= 111.420%		
62) 4-Bromofluorobenzene	12.853	95	234701	48.699	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	= 97.400%		
Target Compounds						Qvalue
16) Acetone	4.430	43	4038m	2.416	ug/l	
22) Diisopropyl ether	6.671	45	12633	0.958	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022624\
Data File : VN081202.D
Acq On : 26 Feb 2024 18:04
Operator : JC\MD
Sample : P1577-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW10A-MW01S-20240221

Quant Time: Feb 27 00:14:34 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/27/2024
Supervised By :Mahesh Dadoda 02/27/2024

