

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022625\
 Data File : VN085871.D
 Acq On : 26 Feb 2025 15:08
 Operator : JC\MD
 Sample : Q1421-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 P001-CLAY-CF02-01

Quant Time: Feb 26 23:02:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	269829	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	498271	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	443623	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	172512	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	184049	52.612	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.220%
35) Dibromofluoromethane	8.165	113	168175	51.580	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.160%
50) Toluene-d8	10.565	98	552142	46.545	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.100%
62) 4-Bromofluorobenzene	12.847	95	180855	46.273	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.540%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	23745	21.452	ug/l	100
20) Methylene Chloride	5.271	84	6218	1.747	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022625\
Data File : VN085871.D
Acq On : 26 Feb 2025 15:08
Operator : JC\MD
Sample : Q1421-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
P001-CLAY-CF02-01

Quant Time: Feb 26 23:02:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

