

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022625\
 Data File : VN085873.D
 Acq On : 26 Feb 2025 15:57
 Operator : JC\MD
 Sample : Q1426-01 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 50453

Quant Time: Feb 26 23:04:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	236480	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	438685	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	394463	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	160274	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	177467	57.884	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	115.760%	
35) Dibromofluoromethane	8.165	113	161536	56.274	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	112.540%	
50) Toluene-d8	10.565	98	535879	51.310	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	102.620%	
62) 4-Bromofluorobenzene	12.847	95	163128	47.407	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	94.820%	
Target Compounds						
16) Acetone	4.448	43	3048	3.142	ug/l	Qvalue # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022625\
Data File : VN085873.D
Acq On : 26 Feb 2025 15:57
Operator : JC\MD
Sample : Q1426-01 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
50453

Quant Time: Feb 26 23:04:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

