

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022625\
 Data File : VN085878.D
 Acq On : 26 Feb 2025 17:57
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 26 23:13:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	0.680	0.598	12.1	80	0.00
3 P	Chloromethane	0.682	0.574	15.8	82	0.00
4 C	Vinyl Chloride	0.723	0.624	13.7#	82	0.00
5 T	Bromomethane	0.462	0.387	16.2	82	0.00
6 T	Chloroethane	0.479	0.393	18.0	84	0.00
7 T	Trichlorofluoromethane	1.046	0.917	12.3	85	0.00
8 T	Diethyl Ether	0.339	0.336	0.9	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.605	0.553	8.6	87	0.00
10 T	Methyl Iodide	0.745	0.706	5.2	87	0.00
11 T	Tert butyl alcohol	0.080	0.087	-8.7	99	0.01
12 CM	1,1-Dichloroethene	0.548	0.517	5.7#	87	0.00
13 T	Acrolein	0.118	0.118	0.0	83	0.00
14 T	Allyl chloride	0.720	0.679	5.7	88	0.00
15 T	Acrylonitrile	0.259	0.270	-4.2	98	0.00
16 T	Acetone	0.205	0.190	7.3	92	0.00
17 T	Carbon Disulfide	1.625	1.307	19.6	79	0.00
18 T	Methyl Acetate	0.705	0.716	-1.6	103	0.00
19 T	Methyl tert-butyl Ether	1.629	1.719	-5.5	92	0.00
20 T	Methylene Chloride	0.660	0.598	9.4	90	0.00
21 T	trans-1,2-Dichloroethene	0.590	0.541	8.3	88	0.00
22 T	Diisopropyl ether	1.627	1.678	-3.1	91	0.00
23 T	Vinyl Acetate	1.114	1.103	1.0	88	0.00
24 P	1,1-Dichloroethane	1.106	1.032	6.7	89	0.00
25 T	2-Butanone	0.302	0.313	-3.6	95	0.00
26 T	2,2-Dichloropropane	0.983	0.865	12.0	82	0.00
27 T	cis-1,2-Dichloroethene	0.685	0.663	3.2	89	0.00
28 T	Bromochloromethane	0.454	0.471	-3.7	94	0.00
29 T	Tetrahydrofuran	0.195	0.211	-8.2	96	0.00
30 C	Chloroform	1.156	1.072	7.3#	89	0.00
31 T	Cyclohexane	0.926	0.792	14.5	82	0.00
32 T	1,1,1-Trichloroethane	1.028	0.946	8.0	87	0.00
33 S	1,2-Dichloroethane-d4	0.648	0.569	12.2	83	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	90	0.00
35 S	Dibromofluoromethane	0.327	0.304	7.0	86	0.00
36 T	1,1-Dichloropropene	0.463	0.440	5.0	86	0.00
37 T	Ethyl Acetate	0.381	0.385	-1.0	92	0.00
38 T	Carbon Tetrachloride	0.550	0.511	7.1	87	0.00
39 T	Methylcyclohexane	0.455	0.440	3.3	80	0.00
40 TM	Benzene	1.484	1.432	3.5	89	0.00
41 T	Methacrylonitrile	0.204	0.212	-3.9	88	0.01
42 TM	1,2-Dichloroethane	0.483	0.438	9.3	86	0.00
43 T	Isopropyl Acetate	0.807	0.652	19.2	92	0.00
44 TM	Trichloroethene	0.364	0.335	8.0	89	0.00
45 C	1,2-Dichloropropane	0.358	0.349	2.5#	91	0.00
46 T	Dibromomethane	0.253	0.242	4.3	90	0.00
47 T	Bromodichloromethane	0.541	0.514	5.0	89	0.00
48 T	Methyl methacrylate	0.283	0.295	-4.2	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022625\
 Data File : VN085878.D
 Acq On : 26 Feb 2025 17:57
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 26 23:13:15 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.005	0.006	-20.0	101	0.00
50 S	Toluene-d8	1.190	1.134	4.7	85	0.00
51 T	4-Methyl-2-Pentanone	0.367	0.399	-8.7	95	0.00
52 CM	Toluene	0.870	0.892	-2.5#	89	0.00
53 T	t-1,3-Dichloropropene	0.500	0.518	-3.6	89	0.00
54 T	cis-1,3-Dichloropropene	0.548	0.566	-3.3	90	0.00
55 T	1,1,2-Trichloroethane	0.346	0.341	1.4	93	0.00
56 T	Ethyl methacrylate	0.434	0.526	-21.2	95	0.00
57 T	1,3-Dichloropropane	0.577	0.577	0.0	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.195	0.244	-25.1#	126	0.00
59 T	2-Hexanone	0.253	0.287	-13.4	94	0.00
60 T	Dibromochloromethane	0.403	0.413	-2.5	92	0.00
61 T	1,2-Dibromoethane	0.327	0.337	-3.1	94	0.00
62 S	4-Bromofluorobenzene	0.392	0.408	-4.1	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	91	0.00
64 T	Tetrachloroethene	0.387	0.351	9.3	89	0.00
65 PM	Chlorobenzene	1.145	1.101	3.8	91	0.00
66 T	1,1,1,2-Tetrachloroethane	0.412	0.399	3.2	92	0.00
67 C	Ethyl Benzene	1.794	1.866	-4.0#	89	0.00
68 T	m/p-Xylenes	0.682	0.735	-7.8	91	0.00
69 T	o-Xylene	0.648	0.705	-8.8	91	0.00
70 T	Styrene	1.059	1.208	-14.1	92	0.00
71 P	Bromoform	0.306	0.315	-2.9	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
73 T	Isopropylbenzene	3.425	3.552	-3.7	90	0.00
74 T	N-amyl acetate	1.208	1.266	-4.8	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.179	1.115	5.4	97	0.00
76 T	1,2,3-Trichloropropane	1.111	1.004	9.6	97	0.00
77 T	Bromobenzene	0.925	0.889	3.9	90	0.00
78 T	n-propylbenzene	3.923	4.102	-4.6	88	0.00
79 T	2-Chlorotoluene	2.621	2.586	1.3	90	0.00
80 T	1,3,5-Trimethylbenzene	2.790	2.996	-7.4	89	0.00
81 T	trans-1,4-Dichloro-2-butene	0.381	0.428	-12.3	98	0.00
82 T	4-Chlorotoluene	2.598	2.617	-0.7	90	0.00
83 T	tert-Butylbenzene	2.346	2.455	-4.6	87	0.00
84 T	1,2,4-Trimethylbenzene	2.761	3.024	-9.5	90	0.00
85 T	sec-Butylbenzene	3.315	3.440	-3.8	86	0.00
86 T	p-Isopropyltoluene	2.665	2.874	-7.8	86	0.00
87 T	1,3-Dichlorobenzene	1.734	1.648	5.0	92	0.00
88 T	1,4-Dichlorobenzene	1.777	1.630	8.3	90	0.00
89 T	n-Butylbenzene	2.354	2.343	0.5	84	0.00
90 T	Hexachloroethane	0.628	0.570	9.2	89	0.00
91 T	1,2-Dichlorobenzene	1.667	1.596	4.3	92	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.216	0.199	7.9	94	0.00
93 T	1,2,4-Trichlorobenzene	0.794	0.787	0.9	89	0.00
94 T	Hexachlorobutadiene	0.483	0.397	17.8	85	0.00
95 T	Naphthalene	2.181	2.393	-9.7	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022625\
Data File : VN085878.D
Acq On : 26 Feb 2025 17:57
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Feb 26 23:13:15 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.798	0.769	3.6	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6