

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022719\
 Data File : VN053982.D
 Acq On : 27 Feb 2019 21:31
 Operator : JC/SP
 Sample : K1353-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-02-022619-C

Quant Time: Feb 28 07:47:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 26 05:37:40 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	629728	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1150281	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1119405	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	426492	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	412540	55.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.78%	
35) Dibromofluoromethane	7.59	113	371358	49.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.84%	
50) Toluene-d8	10.09	98	1467206	52.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.98%	
62) 4-Bromofluorobenzene	12.40	95	493006	53.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.88%	
Target Compounds						
16) Acetone	3.82	43	28863	2.746	ug/l	95
20) Methylene Chloride	4.56	84	72034	9.860	ug/l #	89
43) Isopropyl Acetate	8.17	43	37468	2.905	ug/l #	90
80) 1,3,5-Trimethylbenzene	12.73	105	6886	0.253	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022719\
Data File : VN053982.D
Acq On : 27 Feb 2019 21:31
Operator : JC/SP
Sample : K1353-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PL-02-022619-C

Quant Time: Feb 28 07:47:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 26 05:37:40 2019
Response via : Initial Calibration

