

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022719\
 Data File : VN053989.D
 Acq On : 28 Feb 2019 00:38
 Operator : JC/SP
 Sample : K1650-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TWP-1

Quant Time: Feb 28 08:05:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 28 04:05:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	567028	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1092609	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1051089	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	392216	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	386028	58.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.16%	
35) Dibromofluoromethane	7.59	113	352319	49.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.74%	
50) Toluene-d8	10.09	98	1359650	51.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.38%	
62) 4-Bromofluorobenzene	12.40	95	461194	52.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.26%	
Target Compounds						
30) Chloroform	7.37	83	408608	36.281	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022719\
Data File : VN053989.D
Acq On : 28 Feb 2019 00:38
Operator : JC/SP
Sample : K1650-18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TWP-1

Quant Time: Feb 28 08:05:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 28 04:05:45 2019
Response via : Initial Calibration

