

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN022720\
 Data File : VN060269.D
 Acq On : 27 Feb 2020 14:12
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN022720

Quant Time: Feb 27 16:40:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.458	0.466	-1.7	105	0.00
3 P	Chloromethane	0.798	0.743	6.9	100	0.00
4 C	Vinyl Chloride	0.684	0.657	3.9#	102	0.00
5 T	Bromomethane	0.301	0.290	3.7	106	0.00
6 T	Chloroethane	0.371	0.367	1.1	106	0.00
7 T	Trichlorofluoromethane	0.875	0.872	0.3	105	0.00
8 T	Diethyl Ether	0.330	0.331	-0.3	104	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.495	0.505	-2.0	106	0.00
10 T	Methyl Iodide	0.452	0.530	-17.3	123	0.00
11 T	Tert butyl alcohol	0.095	0.097	-2.1	103	0.00
12 CM	1,1-Dichloroethene	0.525	0.497	5.3#	104	0.00
13 T	Acrolein	0.089	0.091	-2.2	97	0.00
14 T	Allyl chloride	0.886	0.827	6.7	95	0.00
15 T	Acrylonitrile	0.251	0.262	-4.4	103	0.00
16 T	Acetone	0.207	0.187	9.7	93	0.00
17 T	Carbon Disulfide	1.490	1.333	10.5	105	0.00
18 T	Methyl Acetate	0.691	0.664	3.9	104	0.00
19 T	Methyl tert-butyl Ether	1.744	1.764	-1.1	105	0.00
20 T	Methylene Chloride	0.569	0.564	0.9	106	0.00
21 T	trans-1,2-Dichloroethene	0.595	0.527	11.4	104	0.00
22 T	Diisopropyl ether	1.794	1.849	-3.1	106	0.00
23 T	Vinyl Acetate	1.426	1.500	-5.2	105	0.00
24 P	1,1-Dichloroethane	1.016	1.037	-2.1	106	0.00
25 T	2-Butanone	0.335	0.336	-0.3	101	0.00
26 T	2,2-Dichloropropane	0.925	0.954	-3.1	108	0.00
27 T	cis-1,2-Dichloroethene	0.660	0.620	6.1	104	0.00
28 T	Bromochloromethane	0.417	0.354	15.1	94	0.00
29 T	Tetrahydrofuran	0.233	0.230	1.3	101	0.00
30 C	Chloroform	1.011	1.025	-1.4#	105	0.00
31 T	Cyclohexane	0.942	0.909	3.5	105	0.00
32 T	1,1,1-Trichloroethane	0.896	0.920	-2.7	106	0.00
33 S	1,2-Dichloroethane-d4	0.642	0.604	5.9	99	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.232	0.251	-8.2	109	0.00
36 T	1,1-Dichloropropene	0.477	0.485	-1.7	106	0.00
37 T	Ethyl Acetate	0.451	0.471	-4.4	102	0.00
38 T	Carbon Tetrachloride	0.464	0.491	-5.8	106	0.00
39 T	Methylcyclohexane	0.562	0.586	-4.3	108	0.00
40 TM	Benzene	1.387	1.437	-3.6	105	0.00
41 T	Methacrylonitrile	0.208	0.203	2.4	93	0.00
42 TM	1,2-Dichloroethane	0.500	0.521	-4.2	104	0.00
43 T	Isopropyl Acetate	0.790	0.833	-5.4	104	0.00
44 TM	Trichloroethene	0.386	0.395	-2.3	106	0.00
45 C	1,2-Dichloropropane	0.363	0.385	-6.1#	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN022720\
 Data File : VN060269.D
 Acq On : 27 Feb 2020 14:12
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 ICVVN022720

Quant Time: Feb 27 16:40:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 27 13:52:50 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.235	0.246	-4.7	104	0.00
47 T	Bromodichloromethane	0.479	0.513	-7.1	105	0.00
48 T	Methyl methacrylate	0.360	0.373	-3.6	103	0.00
49 T	1,4-Dioxane	0.005	0.005	0.0	110	0.00
50 S	Toluene-d8	1.175	1.143	2.7	101	0.00
51 T	4-Methyl-2-Pentanone	0.448	0.478	-6.7	103	0.00
52 CM	Toluene	0.871	0.915	-5.1#	106	0.00
53 T	t-1,3-Dichloropropene	0.586	0.605	-3.2	105	0.00
54 T	cis-1,3-Dichloropropene	0.608	0.639	-5.1	106	0.00
55 T	1,1,2-Trichloroethane	0.350	0.359	-2.6	105	0.00
56 T	Ethyl methacrylate	0.503	0.560	-11.3	105	0.00
57 T	1,3-Dichloropropane	0.571	0.600	-5.1	104	0.00
58 T	2-Chloroethyl Vinyl ether	0.263	0.276	-4.9	102	0.00
59 T	2-Hexanone	0.329	0.349	-6.1	102	0.00
60 T	Dibromochloromethane	0.356	0.392	-10.1	108	0.00
61 T	1,2-Dibromoethane	0.342	0.364	-6.4	105	0.00
62 S	4-Bromofluorobenzene	0.445	0.431	3.1	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	98	0.00
64 T	Tetrachloroethene	0.444	0.443	0.2	107	0.00
65 PM	Chlorobenzene	1.010	1.040	-3.0	105	0.00
66 T	1,1,1,2-Tetrachloroethane	0.373	0.400	-7.2	107	0.00
67 C	Ethyl Benzene	1.839	1.954	-6.3#	107	0.00
68 T	m/p-Xylenes	0.694	0.726	-4.6	107	0.00
69 T	o-Xylene	0.661	0.704	-6.5	108	0.00
70 T	Styrene	1.088	1.178	-8.3	107	0.00
71 P	Bromoform	0.269	0.300	-11.5	105	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.706	3.910	-5.5	107	0.00
74 T	N-amyl acetate	1.619	1.667	-3.0	104	0.00
75 P	1,1,2,2-Tetrachloroethane	1.007	1.043	-3.6	103	0.00
76 T	1,2,3-Trichloropropane	1.097	1.036	5.6	95	0.00
77 T	Bromobenzene	0.908	0.947	-4.3	107	0.00
78 T	n-propylbenzene	4.344	4.575	-5.3	107	0.00
79 T	2-Chlorotoluene	2.554	2.646	-3.6	106	0.00
80 T	1,3,5-Trimethylbenzene	3.120	3.326	-6.6	107	0.00
81 T	trans-1,4-Dichloro-2-butene	0.397	0.443	-11.6	108	0.00
82 T	4-Chlorotoluene	2.690	2.777	-3.2	107	0.00
83 T	tert-Butylbenzene	2.701	2.865	-6.1	107	0.00
84 T	1,2,4-Trimethylbenzene	3.149	3.321	-5.5	106	0.00
85 T	sec-Butylbenzene	3.618	3.846	-6.3	107	0.00
86 T	p-Isopropyltoluene	3.348	3.510	-4.8	108	0.00
87 T	1,3-Dichlorobenzene	1.706	1.725	-1.1	106	0.00
88 T	1,4-Dichlorobenzene	1.697	1.701	-0.2	106	0.00
89 T	n-Butylbenzene	3.074	3.187	-3.7	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN022720\
Data File : VN060269.D
Acq On : 27 Feb 2020 14:12
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN022720

Quant Time: Feb 27 16:40:59 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022720W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 27 13:52:50 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.485	0.539	-11.1	109	0.00
91 T	1,2-Dichlorobenzene	1.592	1.631	-2.4	105	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.228	0.224	1.8	99	0.00
93 T	1,2,4-Trichlorobenzene	1.130	1.059	6.3	103	0.00
94 T	Hexachlorobutadiene	0.536	0.513	4.3	107	0.00
95 T	Naphthalene	3.258	3.060	6.1	99	0.00
96 T	1,2,3-Trichlorobenzene	1.062	0.997	6.1	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6