

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022723\
 Data File : VN076820.D
 Acq On : 27 Feb 2023 12:55
 Operator : JC\MD
 Sample : 01653-01 2500X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-0567

Quant Time: Feb 28 01:50:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.231	168	257332	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.107	114	474686	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.872	117	407150	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.795	152	160159	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	186577	52.507	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.020%	
35) Dibromofluoromethane	8.172	113	145980	46.960	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 93.920%	
50) Toluene-d8	10.572	98	571823	49.255	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.520%	
62) 4-Bromofluorobenzene	12.854	95	162477	42.256	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 84.520%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022723\
Data File : VN076820.D
Acq On : 27 Feb 2023 12:55
Operator : JC\MD
Sample : 01653-01 2500X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OK-0567

Quant Time: Feb 28 01:50:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

