

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022723\
 Data File : VN076814.D
 Acq On : 27 Feb 2023 09:42
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 28 01:49:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	77	0.00
2 T	Dichlorodifluoromethane	0.531	0.434	18.3	61	0.00
3 P	Chloromethane	0.693	0.560	19.2	70	0.00
4 C	Vinyl Chloride	0.682	0.588	13.8#	69	0.00
5 T	Bromomethane	0.413	0.364	11.9	79	0.00
6 T	Chloroethane	0.431	0.429	0.5	78	0.00
7 T	Trichlorofluoromethane	0.928	0.797	14.1	71	0.00
8 T	Diethyl Ether	0.357	0.335	6.2	73	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.545	0.506	7.2	78	0.00
10 T	Methyl Iodide	0.673	0.649	3.6	66	0.00
11 T	Tert butyl alcohol	0.091	0.086	5.5	75	0.00
12 CM	1,1-Dichloroethene	0.516	0.444	14.0#	71	0.00
13 T	Acrolein	0.137	0.102	25.5#	60	0.00
14 T	Allyl chloride	0.894	0.844	5.6	76	0.00
15 T	Acrylonitrile	0.323	0.308	4.6	76	0.00
16 T	Acetone	0.308	0.328	-6.5	95	0.00
17 T	Carbon Disulfide	1.363	1.092	19.9	65	0.00
18 T	Methyl Acetate	1.007	0.969	3.8	77	0.00
19 T	Methyl tert-butyl Ether	1.740	1.772	-1.8	78	0.00
20 T	Methylene Chloride	0.646	0.567	12.2	73	0.00
21 T	trans-1,2-Dichloroethene	0.581	0.511	12.0	74	0.00
22 T	Diisopropyl ether	2.097	2.268	-8.2	83	0.00
23 T	Vinyl Acetate	1.241	1.434	-15.6	81	0.00
24 P	1,1-Dichloroethane	1.149	1.111	3.3	80	0.00
25 T	2-Butanone	0.455	0.460	-1.1	82	0.00
26 T	2,2-Dichloropropane	0.829	0.861	-3.9	83	0.00
27 T	cis-1,2-Dichloroethene	0.689	0.656	4.8	77	0.00
28 T	Bromochloromethane	0.547	0.622	-13.7	91	0.00
29 T	Tetrahydrofuran	0.297	0.300	-1.0	78	0.00
30 C	Chloroform	1.148	1.142	0.5#	81	0.00
31 T	Cyclohexane	1.083	0.998	7.8	76	0.00
32 T	1,1,1-Trichloroethane	0.966	0.955	1.1	79	0.00
33 S	1,2-Dichloroethane-d4	0.690	0.740	-7.2	87	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	79	0.00
35 S	Dibromofluoromethane	0.327	0.330	-0.9	84	0.00
36 T	1,1-Dichloropropene	0.473	0.473	0.0	79	0.00
37 T	Ethyl Acetate	0.498	0.524	-5.2	81	0.00
38 T	Carbon Tetrachloride	0.477	0.467	2.1	78	0.00
39 T	Methylcyclohexane	0.538	0.559	-3.9	77	0.00
40 TM	Benzene	1.453	1.481	-1.9	80	0.00
41 T	Methacrylonitrile	0.288	0.289	-0.3	76	0.00
42 TM	1,2-Dichloroethane	0.507	0.522	-3.0	82	0.00
43 T	Isopropyl Acetate	0.810	0.872	-7.7	79	0.00
44 TM	Trichloroethene	0.343	0.332	3.2	75	0.00
45 C	1,2-Dichloropropane	0.385	0.400	-3.9#	82	0.00
46 T	Dibromomethane	0.246	0.259	-5.3	83	0.00
47 T	Bromodichloromethane	0.493	0.535	-8.5	85	0.00
48 T	Methyl methacrylate	0.384	0.422	-9.9	84	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022723\
 Data File : VN076814.D
 Acq On : 27 Feb 2023 09:42
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 28 01:49:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Feb 14 01:44:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.007	0.007	0.0	80	0.00
50 S	Toluene-d8	1.223	1.342	-9.7	87	0.00
51 T	4-Methyl-2-Pentanone	0.515	0.558	-8.3	82	0.00
52 CM	Toluene	0.877	0.962	-9.7#	84	0.00
53 T	t-1,3-Dichloropropene	0.476	0.564	-18.5	86	0.00
54 T	cis-1,3-Dichloropropene	0.549	0.612	-11.5	84	0.00
55 T	1,1,2-Trichloroethane	0.357	0.383	-7.3	85	0.00
56 T	Ethyl methacrylate	0.502	0.572	-13.9	80	0.00
57 T	1,3-Dichloropropane	0.613	0.664	-8.3	85	0.00
58 T	2-Chloroethyl Vinyl ether	0.230	0.246	-7.0	75	0.00
59 T	2-Hexanone	0.372	0.418	-12.4	84	0.00
60 T	Dibromochloromethane	0.376	0.414	-10.1	84	0.00
61 T	1,2-Dibromoethane	0.357	0.376	-5.3	83	0.00
62 S	4-Bromofluorobenzene	0.405	0.486	-20.0	92	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	83	0.00
64 T	Tetrachloroethene	0.315	0.292	7.3	76	0.00
65 PM	Chlorobenzene	1.060	1.043	1.6	83	0.00
66 T	1,1,1,2-Tetrachloroethane	0.385	0.397	-3.1	83	0.00
67 C	Ethyl Benzene	1.805	1.940	-7.5#	86	0.00
68 T	m/p-Xylenes	0.696	0.757	-8.8	84	0.00
69 T	o-Xylene	0.679	0.750	-10.5	86	0.00
70 T	Styrene	1.097	1.275	-16.2	88	0.00
71 P	Bromoform	0.282	0.300	-6.4	84	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
73 T	Isopropylbenzene	3.665	3.664	0.0	87	0.00
74 T	N-amyl acetate	1.652	1.533	7.2	79	0.00
75 P	1,1,2,2-Tetrachloroethane	1.327	1.165	12.2	87	0.00
76 T	1,2,3-Trichloropropane	1.011	0.785	22.4	72	0.00
77 T	Bromobenzene	0.914	0.831	9.1	84	0.00
78 T	n-propylbenzene	4.303	4.517	-5.0	90	0.00
79 T	2-Chlorotoluene	2.635	2.583	2.0	89	0.00
80 T	1,3,5-Trimethylbenzene	3.029	3.121	-3.0	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.347	0.353	-1.7	86	0.00
82 T	4-Chlorotoluene	2.635	2.583	2.0	89	0.00
83 T	tert-Butylbenzene	2.740	2.780	-1.5	87	0.00
84 T	1,2,4-Trimethylbenzene	3.019	3.205	-6.2	90	0.00
85 T	sec-Butylbenzene	3.825	4.106	-7.3	89	0.00
86 T	p-Isopropyltoluene	3.055	3.405	-11.5	90	0.00
87 T	1,3-Dichlorobenzene	1.728	1.711	1.0	89	0.00
88 T	1,4-Dichlorobenzene	1.721	1.686	2.0	89	0.00
89 T	n-Butylbenzene	2.657	2.971	-11.8	89	0.00
90 T	Hexachloroethane	0.640	0.631	1.4	91	0.00
91 T	1,2-Dichlorobenzene	1.772	1.691	4.6	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.257	0.211	17.9	81	0.00
93 T	1,2,4-Trichlorobenzene	0.884	0.856	3.2	79	0.00
94 T	Hexachlorobutadiene	0.433	0.400	7.6	81	0.00
95 T	Naphthalene	2.682	2.598	3.1	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022723\
Data File : VN076814.D
Acq On : 27 Feb 2023 09:42
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Feb 28 01:49:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021323W.M
Quant Title : SW846 8260
QLast Update : Tue Feb 14 01:44:47 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.856	0.821	4.1	77	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6