

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
 Data File : VN081209.D
 Acq On : 27 Feb 2024 10:53
 Operator : JC\MD
 Sample : VN0227MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0227MBL01

Quant Time: Feb 28 04:13:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	355778	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	651894	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	558392	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	223485	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	238442	54.678	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.360%
35) Dibromofluoromethane	8.171	113	187057	50.744	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.480%
50) Toluene-d8	10.570	98	719862	54.982	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	109.960%
62) 4-Bromofluorobenzene	12.853	95	237721	48.321	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
Data File : VN081209.D
Acq On : 27 Feb 2024 10:53
Operator : JC\MD
Sample : VN0227MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0227MBL01

Quant Time: Feb 28 04:13:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

