

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
 Data File : VN081217.D
 Acq On : 27 Feb 2024 14:22
 Operator : JC\MD
 Sample : P1577-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP07-20240221

Quant Time: Feb 28 04:17:53 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	339568	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	632619	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	544759	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	220825	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	227040	54.549	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.100%	
35) Dibromofluoromethane	8.171	113	178845	49.995	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.980%	
50) Toluene-d8	10.571	98	677353	53.312	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.620%	
62) 4-Bromofluorobenzene	12.853	95	227420	47.636	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.280%	
Target Compounds						
22) Diisopropyl ether	6.671	45	12414	0.960	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
Data File : VN081217.D
Acq On : 27 Feb 2024 14:22
Operator : JC\MD
Sample : P1577-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP07-20240221

Quant Time: Feb 28 04:17:53 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

