

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
 Data File : VN081218.D
 Acq On : 27 Feb 2024 14:46
 Operator : JC\MD
 Sample : P1585-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-232

Quant Time: Feb 28 04:18:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	323698	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	616571	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	555749	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	193066	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	216226	54.498	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.000%
35) Dibromofluoromethane	8.171	113	173529	49.771	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.540%
50) Toluene-d8	10.570	98	688787	55.623	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	111.240%
62) 4-Bromofluorobenzene	12.853	95	211790	45.517	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.040%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	87191	55.850	ug/l	99
20) Methylene Chloride	5.283	84	24488	5.231	ug/l	93
43) Isopropyl Acetate	8.688	43	515391	54.665	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
Data File : VN081218.D
Acq On : 27 Feb 2024 14:46
Operator : JC\MD
Sample : P1585-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-232

Quant Time: Feb 28 04:18:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

