

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
 Data File : VN081219.D
 Acq On : 27 Feb 2024 15:11
 Operator : JC\MD
 Sample : P1577-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RW10-MW01S-20240221

Quant Time: Feb 28 04:18:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	326740	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	614642	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	521087	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	199772	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	217048	54.195	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 108.400%		
35) Dibromofluoromethane	8.171	113	175421	50.472	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.940%		
50) Toluene-d8	10.570	98	666279	53.974	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 107.940%		
62) 4-Bromofluorobenzene	12.853	95	214601	46.266	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 92.540%		
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	4375	2.776	ug/l #	86
19) Methyl tert-butyl Ether	5.800	73	11442	0.904	ug/l #	82
27) cis-1,2-Dichloroethene	7.488	96	3804	0.794	ug/l	72
64) Tetrachloroethene	11.106	164	13608	3.826	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
Data File : VN081219.D
Acq On : 27 Feb 2024 15:11
Operator : JC\MD
Sample : P1577-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RW10-MW01S-20240221

Quant Time: Feb 28 04:18:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

