

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
 Data File : VN081220.D
 Acq On : 27 Feb 2024 15:35
 Operator : JC\MD
 Sample : P1590-02 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE120D1-20240222

Quant Time: Feb 28 04:19:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 02/28/2024
 Supervised By : Semsettin Yesilyurt 02/28/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	316987	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	599183	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	522958	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	202275	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	213012	54.824	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 109.640%			
35) Dibromofluoromethane	8.171	113	168837	49.831	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.660%			
50) Toluene-d8	10.571	98	654742	54.408	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 108.820%			
62) 4-Bromofluorobenzene	12.853	95	208063	46.013	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 92.020%			
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	4.342	96	2215m	0.618	ug/l	
27) cis-1,2-Dichloroethene	7.500	96	1356	0.292	ug/l	75
44) Trichloroethene	9.359	130	244847	54.714	ug/l	99
64) Tetrachloroethene	11.112	164	2401	0.673	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
Data File : VN081220.D
Acq On : 27 Feb 2024 15:35
Operator : JC\MD
Sample : P1590-02 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE120D1-20240222

Quant Time: Feb 28 04:19:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/28/2024
Supervised By :Semsettin Yesilyurt 02/28/2024

