

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
 Data File : VN081225.D
 Acq On : 27 Feb 2024 17:36
 Operator : JC\MD
 Sample : P1590-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB06-20240222

Quant Time: Feb 28 04:21:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	314045	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	588686	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	518438	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	198891	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	214567	55.742	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.480%
35) Dibromofluoromethane	8.165	113	170873	51.331	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.660%
50) Toluene-d8	10.571	98	653331	55.259	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	110.520%
62) 4-Bromofluorobenzene	12.853	95	208858	47.013	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
Data File : VN081225.D
Acq On : 27 Feb 2024 17:36
Operator : JC\MD
Sample : P1590-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB06-20240222

Quant Time: Feb 28 04:21:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

