

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
 Data File : VN081226.D
 Acq On : 27 Feb 2024 18:00
 Operator : JC\MD
 Sample : P1590-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT158S1-20240222

Quant Time: Feb 28 04:22:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	307864	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	570630	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	509779	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	198349	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	212306	56.262	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 112.520%	
35) Dibromofluoromethane	8.171	113	167860	52.022	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.040%	
50) Toluene-d8	10.571	98	639906	55.836	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 111.680%	
62) 4-Bromofluorobenzene	12.853	95	204402	47.466	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.940%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	5015	3.378	ug/l #	86
24) 1,1-Dichloroethane	6.571	63	11430	1.687	ug/l #	85
52) Toluene	10.635	92	8959	0.884	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
Data File : VN081226.D
Acq On : 27 Feb 2024 18:00
Operator : JC\MD
Sample : P1590-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT158S1-20240222

Quant Time: Feb 28 04:22:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

