

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
 Data File : VN081232.D
 Acq On : 27 Feb 2024 20:24
 Operator : JC\MD
 Sample : P1590-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BPOW1-7-20240223

Manual Integrations
 APPROVED

Reviewed By : John Carlone 02/28/2024
 Supervised By : Semsettin Yesilyurt 02/28/2024

Quant Time: Feb 28 04:24:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	310282	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	583122	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	506724	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	199124	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	215314	56.614	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 113.220%			
35) Dibromofluoromethane	8.171	113	167904	50.920	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.840%			
50) Toluene-d8	10.571	98	636663	54.363	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 108.720%			
62) 4-Bromofluorobenzene	12.853	95	204085	46.377	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 92.760%			
Target Compounds						
16) Acetone	4.436	43	5296m	3.539	ug/l	
52) Toluene	10.635	92	18261	1.764	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
Data File : VN081232.D
Acq On : 27 Feb 2024 20:24
Operator : JC\MD
Sample : P1590-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BPOW1-7-20240223

Quant Time: Feb 28 04:24:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 02/28/2024
Supervised By :Semsettin Yesilyurt 02/28/2024

