

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
 Data File : VN081233.D
 Acq On : 27 Feb 2024 20:48
 Operator : JC\MD
 Sample : P1590-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE120D3-20240222

Quant Time: Feb 28 04:25:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	310148	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	575259	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	499241	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	191493	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	216502	56.951	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.900%
35) Dibromofluoromethane	8.171	113	171541	52.735	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.460%
50) Toluene-d8	10.571	98	643977	55.739	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	111.480%
62) 4-Bromofluorobenzene	12.853	95	203098	46.783	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.560%
Target Compounds						
						Qvalue
16) Acetone	4.442	43	6004	4.014	ug/l	90
44) Trichloroethene	9.353	130	11309	2.632	ug/l	79
52) Toluene	10.629	92	8348	0.817	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
 Data File : VN081233.D
 Acq On : 27 Feb 2024 20:48
 Operator : JC\MD
 Sample : P1590-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE120D3-20240222

Quant Time: Feb 28 04:25:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

