

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
 Data File : VN081208.D
 Acq On : 27 Feb 2024 10:18
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 28 04:12:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	73	0.00
2 T	Dichlorodifluoromethane	0.647	0.649	-0.3	69	0.00
3 P	Chloromethane	0.833	0.646	22.4	66	0.00
4 C	Vinyl Chloride	0.718	0.709	1.3#	73	0.00
5 T	Bromomethane	0.510	0.443	13.1	66	0.00
6 T	Chloroethane	0.556	0.490	11.9	77	0.00
7 T	Trichlorofluoromethane	1.049	1.082	-3.1	76	0.00
8 T	Diethyl Ether	0.384	0.403	-4.9	77	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.585	0.617	-5.5	78	0.01
10 T	Methyl Iodide	0.653	0.606	7.2	65	0.00
11 T	Tert butyl alcohol	0.113	0.107	5.3	76	0.00
12 CM	1,1-Dichloroethene	0.565	0.551	2.5#	72	0.00
13 T	Acrolein	0.060	0.079	-31.7#	104	0.00
14 T	Allyl chloride	0.839	0.848	-1.1	75	0.00
15 T	Acrylonitrile	0.294	0.328	-11.6	84	0.00
16 T	Acetone	0.241	0.228	5.4	74	0.00
17 T	Carbon Disulfide	1.577	1.413	10.4	68	0.00
18 T	Methyl Acetate	0.588	0.765	-30.1#	94	0.00
19 T	Methyl tert-butyl Ether	1.937	2.110	-8.9	80	0.00
20 T	Methylene Chloride	0.693	0.684	1.3	80	0.00
21 T	trans-1,2-Dichloroethene	0.631	0.631	0.0	77	0.00
22 T	Diisopropyl ether	1.903	2.107	-10.7	81	0.00
23 T	Vinyl Acetate	1.544	1.676	-8.5	78	0.00
24 P	1,1-Dichloroethane	1.100	1.187	-7.9	79	0.00
25 T	2-Butanone	0.390	0.410	-5.1	81	0.00
26 T	2,2-Dichloropropane	1.031	1.064	-3.2	75	0.00
27 T	cis-1,2-Dichloroethene	0.733	0.765	-4.4	77	0.00
28 T	Bromochloromethane	0.456	0.555	-21.7	90	0.00
29 T	Tetrahydrofuran	0.260	0.282	-8.5	83	0.00
30 C	Chloroform	1.170	1.307	-11.7#	83	0.00
31 T	Cyclohexane	1.058	1.008	4.7	76	0.00
32 T	1,1,1-Trichloroethane	1.065	1.105	-3.8	77	0.00
33 S	1,2-Dichloroethane-d4	0.613	0.642	-4.7	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	73	0.00
35 S	Dibromofluoromethane	0.283	0.306	-8.1	81	0.00
36 T	1,1-Dichloropropene	0.484	0.489	-1.0	76	0.00
37 T	Ethyl Acetate	0.437	0.471	-7.8	80	0.00
38 T	Carbon Tetrachloride	0.490	0.530	-8.2	77	0.00
39 T	Methylcyclohexane	0.565	0.576	-1.9	74	0.00
40 TM	Benzene	1.402	1.501	-7.1	78	0.00
41 T	Methacrylonitrile	0.249	0.270	-8.4	88	0.00
42 TM	1,2-Dichloroethane	0.489	0.543	-11.0	83	0.00
43 T	Isopropyl Acetate	0.765	0.825	-7.8	80	0.00
44 TM	Trichloroethene	0.373	0.378	-1.3	77	0.00
45 C	1,2-Dichloropropane	0.351	0.390	-11.1#	82	0.00
46 T	Dibromomethane	0.256	0.279	-9.0	82	0.00
47 T	Bromodichloromethane	0.484	0.561	-15.9	82	0.00
48 T	Methyl methacrylate	0.356	0.382	-7.3	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
 Data File : VN081208.D
 Acq On : 27 Feb 2024 10:18
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Feb 28 04:12:18 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
 Quant Title : SW846 8260
 QLast Update : Fri Feb 16 23:45:20 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	83	0.00
50 S	Toluene-d8	1.004	0.989	1.5	76	0.00
51 T	4-Methyl-2-Pentanone	0.440	0.502	-14.1	84	0.00
52 CM	Toluene	0.888	0.959	-8.0#	78	0.00
53 T	t-1,3-Dichloropropene	0.538	0.606	-12.6	77	0.00
54 T	cis-1,3-Dichloropropene	0.581	0.637	-9.6	78	0.00
55 T	1,1,2-Trichloroethane	0.341	0.388	-13.8	81	0.00
56 T	Ethyl methacrylate	0.522	0.585	-12.1	79	0.00
57 T	1,3-Dichloropropane	0.590	0.664	-12.5	82	0.00
58 T	2-Chloroethyl Vinyl ether	0.262	0.287	-9.5	80	0.00
59 T	2-Hexanone	0.349	0.375	-7.4	74	0.00
60 T	Dibromochloromethane	0.363	0.410	-12.9	77	0.00
61 T	1,2-Dibromoethane	0.359	0.399	-11.1	80	0.00
62 S	4-Bromofluorobenzene	0.377	0.398	-5.6	80	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	74	0.00
64 T	Tetrachloroethene	0.341	0.391	-14.7	91	0.00
65 PM	Chlorobenzene	1.082	1.121	-3.6	79	0.00
66 T	1,1,1,2-Tetrachloroethane	0.389	0.431	-10.8	81	0.00
67 C	Ethyl Benzene	1.936	2.032	-5.0#	77	0.00
68 T	m/p-Xylenes	0.742	0.769	-3.6	76	0.00
69 T	o-Xylene	0.736	0.753	-2.3	76	0.00
70 T	Styrene	1.166	1.283	-10.0	79	0.00
71 P	Bromoform	0.247	0.289	-17.0	80	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
73 T	Isopropylbenzene	4.375	4.262	2.6	79	0.00
74 T	N-amyl acetate	1.849	1.758	4.9	76	0.00
75 P	1,1,2,2-Tetrachloroethane	1.230	1.229	0.1	84	0.00
76 T	1,2,3-Trichloropropane	1.136	1.150	-1.2	88	0.00
77 T	Bromobenzene	0.975	0.944	3.2	81	0.00
78 T	n-propylbenzene	5.036	5.098	-1.2	81	0.00
79 T	2-Chlorotoluene	3.089	3.014	2.4	82	0.00
80 T	1,3,5-Trimethylbenzene	3.547	3.592	-1.3	80	0.00
81 T	trans-1,4-Dichloro-2-butene	0.441	0.435	1.4	79	0.00
82 T	4-Chlorotoluene	3.020	2.962	1.9	81	0.00
83 T	tert-Butylbenzene	3.067	3.004	2.1	79	0.00
84 T	1,2,4-Trimethylbenzene	3.534	3.615	-2.3	82	0.00
85 T	sec-Butylbenzene	4.264	4.472	-4.9	83	0.00
86 T	p-Isopropyltoluene	3.542	3.687	-4.1	83	0.00
87 T	1,3-Dichlorobenzene	1.843	1.775	3.7	81	0.00
88 T	1,4-Dichlorobenzene	1.811	1.804	0.4	83	0.00
89 T	n-Butylbenzene	3.085	3.298	-6.9	84	0.00
90 T	Hexachloroethane	0.585	0.617	-5.5	80	0.00
91 T	1,2-Dichlorobenzene	1.778	1.717	3.4	81	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.256	0.248	3.1	81	0.00
93 T	1,2,4-Trichlorobenzene	0.992	1.000	-0.8	81	0.00
94 T	Hexachlorobutadiene	0.370	0.359	3.0	79	0.00
95 T	Naphthalene	3.634	3.546	2.4	79	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022724\
Data File : VN081208.D
Acq On : 27 Feb 2024 10:18
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Feb 28 04:12:18 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021624W.M
Quant Title : SW846 8260
QLast Update : Fri Feb 16 23:45:20 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.984	0.972	1.2	79	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6