

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022725\
 Data File : VN085888.D
 Acq On : 27 Feb 2025 14:28
 Operator : JC\MD
 Sample : PB166880ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166880ZHE#01

Quant Time: Feb 28 02:00:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	8.224	168	207564	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	398078	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	371093	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	134346	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	149869	55.693	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 111.380%			
35) Dibromofluoromethane	8.165	113	135307	51.945	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.880%			
50) Toluene-d8	10.565	98	509258	53.735	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 107.480%			
62) 4-Bromofluorobenzene	12.847	95	155365	49.756	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 99.520%			
Target Compounds						Qvalue
3) Chloromethane	2.359	50	677	0.239	ug/l #	41
5) Bromomethane	2.983	94	557	0.290	ug/l #	49
11) Tert butyl alcohol	5.512	59	151	0.456	ug/l #	72
16) Acetone	4.436	43	48189	56.594	ug/l	95
18) Methyl Acetate	5.042	43	2757	0.941	ug/l #	83
20) Methylene Chloride	5.271	84	4382	1.600	ug/l	90
25) 2-Butanone	7.483	43	5167	4.125	ug/l	96
30) Chloroform	7.959	83	2958	0.617	ug/l	84
31) Cyclohexane	8.224	56	4286	1.115	ug/l #	1
36) 1,1-Dichloropropene	8.224	75	16170	4.386	ug/l #	50
37) Ethyl Acetate	7.559	43	4395	1.449	ug/l #	86
38) Carbon Tetrachloride	8.218	117	21817	4.981	ug/l #	19
42) 1,2-Dichloroethane	8.694	62	1251	0.325	ug/l #	73
43) Isopropyl Acetate	8.688	43	175654	32.803	ug/l #	90
45) 1,2-Dichloropropane	9.747	63	2421	0.849	ug/l #	43
48) Methyl methacrylate	9.741	41	146346	64.954	ug/l #	35
49) 1,4-Dioxane	9.735	88	55	1.284	ug/l #	1
51) 4-Methyl-2-Pentanone	10.447	43	12448	4.256	ug/l #	76
53) t-1,3-Dichloropropene	10.718	75	1372	0.345	ug/l #	46
54) cis-1,3-Dichloropropene	10.359	75	66583	15.257	ug/l #	59
56) Ethyl methacrylate	10.977	69	2482	1.159	ug/l #	1
57) 1,3-Dichloropropane	11.200	76	9271	2.017	ug/l #	42
59) 2-Hexanone	11.200	43	93271	46.352	ug/l #	56
70) Styrene	12.412	104	64	0.592	ug/l #	36
74) N-ethyl acetate	12.482	43	1611	0.496	ug/l #	42
76) 1,2,3-Trichloropropane	12.847	75	72547	24.310	ug/l #	1
81) trans-1,4-Dichloro-2-b...	12.847	75	72547	70.930	ug/l #	12
84) 1,2,4-Trimethylbenzene	13.482	105	1579	0.213	ug/l	93
86) p-Isopropyltoluene	13.953	119	59	0.633	ug/l #	50
95) Naphthalene	15.647	128	2121	0.362	ug/l #	82


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022725\
Data File : VN085888.D
Acq On : 27 Feb 2025 14:28
Operator : JC\MD
Sample : PB166880ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166880ZHE#01

Quant Time: Feb 28 02:00:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28




Instrument :
MSVOA_N
ClientSampleId :
PB166880ZHE#01


Tgt Ion:168 Resp: 207564
Ion Ratio Lower Upper
168 100
99 58.2 47.9 71.9



#3
Chloromethane
Concen: 0.239 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28

Tgt Ion: 50 Resp: 677
Ion Ratio Lower Upper
50 100
52 0.0 26.7 40.1#









#5
Bromomethane
Concen: 0.290 ug/l
RT: 2.983 min Scan# 11
Delta R.T. 0.029 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28




Instrument :
MSVOA_N
ClientSampleId :
PB166880ZHE#01


Tgt Ion: 94 Resp: 557
Ion Ratio Lower Upper
94 100
96 41.3 71.7 107.5#



#11
Tert butyl alcohol
Concen: 0.456 ug/l
RT: 5.512 min Scan# 605
Delta R.T. 0.006 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28

Tgt Ion: 59 Resp: 151
Ion Ratio Lower Upper
59 100
57 0.0 8.3 12.5#









#16
Acetone
Concen: 56.594 ug/l
RT: 4.436 min Scan# 419
Delta R.T. 0.018 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28




Instrument :
MSVOA_N
ClientSampleId :
PB166880ZHE#01


Tgt Ion: 43 Resp: 48189
Ion Ratio Lower Upper
43 100
58 37.3 27.5 41.3



#18
Methyl Acetate
Concen: 0.941 ug/l
RT: 5.042 min Scan# 525
Delta R.T. 0.024 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28

Tgt Ion: 43 Resp: 2757
Ion Ratio Lower Upper
43 100
74 18.8 22.2 33.4#









#20
Methylene Chloride
Concen: 1.600 ug/l
RT: 5.271 min Scan# 51
Delta R.T. 0.000 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28




Instrument :
MSVOA_N
ClientSampleId :
PB166880ZHE#01


Tgt Ion:	84	Resp:	4382
Ion	Ratio	Lower	Upper
84	100		
49	95.0	88.4	132.6
51	31.6	26.6	40.0
86	58.4	51.0	76.6



#25
2-Butanone
Concen: 4.125 ug/l
RT: 7.483 min Scan# 940
Delta R.T. 0.012 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28

Tgt Ion:	43	Resp:	5167
Ion	Ratio	Lower	Upper
43	100		
72	27.4	23.6	35.4





#30

Chloroform

Concen: 0.617 ug/l

RT: 7.959 min Scan# 1021

Delta R.T. 0.000 min

Lab File: VN085888.D

Acq: 27 Feb 2025 14:28

Instrument :

MSVOA_N

ClientSampleId :





PB166880ZHE#01

Tgt Ion: 83 Resp: 2958

Ion Ratio Lower Upper

83 100

85 51.0 50.9 76.3



#31

Cyclohexane

Concen: 1.115 ug/l

RT: 8.224 min Scan# 1066

Delta R.T. -0.029 min

Lab File: VN085888.D

Acq: 27 Feb 2025 14:28




Tgt Ion: 56 Resp: 4286


Ion Ratio Lower Upper

56 100

69 195.6 25.8 38.6#

84 134.6 74.5 111.7#





#33

1,2-Dichloroethane-d4

Concen: 55.693 ug/l

RT: 8.577 min Scan# 1125

Delta R.T. 0.006 min

Lab File: VN085888.D

Acq: 27 Feb 2025 14:28

Instrument :

MSVOA_N

ClientSampleId :





PB166880ZHE#01

Tgt Ion: 65 Resp: 149869

Ion Ratio Lower Upper

65 100

67 53.7 0.0 106.2



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1215

Delta R.T. 0.006 min

Lab File: VN085888.D

Acq: 27 Feb 2025 14:28




Tgt Ion: 114 Resp: 398078


Ion Ratio Lower Upper

114 100

63 19.5 0.0 39.2

88 15.9 0.0 30.0





#35

Dibromofluoromethane

Concen: 51.945 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN085888.D

Acq: 27 Feb 2025 14:28

Instrument :

MSVOA_N

ClientSampleId :

PB166880ZHE#01





Tgt Ion:113 Resp: 135307

Ion Ratio Lower Upper

113 100

111 102.3 82.2 123.4

192 19.3 14.7 22.1



#36

1,1-Dichloropropene

Concen: 4.386 ug/l

RT: 8.224 min Scan# 1066

Delta R.T. -0.141 min

Lab File: VN085888.D

Acq: 27 Feb 2025 14:28




Tgt Ion: 75 Resp: 16170


Ion Ratio Lower Upper

75 100

110 9.1 17.3 52.0#

77 0.0 24.6 37.0#





#37

Ethyl Acetate

Concen: 1.449 ug/l

RT: 7.559 min Scan# 91

Delta R.T. 0.006 min

Lab File: VN085888.D


Acq: 27 Feb 2025 14:28

Instrument :




MSVOA_N

ClientSampleId :

PB166880ZHE#01



Tgt Ion:	43	Resp:	4395
Ion	Ratio	Lower	Upper
43	100		
61	9.0	12.2	18.2#
70	7.0	9.5	14.3#



#38

Carbon Tetrachloride

Concen: 4.981 ug/l




RT: 8.218 min Scan# 1065


Delta R.T. -0.135 min

Lab File: VN085888.D

Acq: 27 Feb 2025 14:28

Tgt Ion:	117	Resp:	21817
Ion	Ratio	Lower	Upper
117	100		
119	7.3	73.8	110.6#
121	0.0	24.3	36.5#





#42

1,2-Dichloroethane

Concen: 0.325 ug/l

RT: 8.694 min Scan# 1146

Delta R.T. 0.030 min

Lab File: VN085888.D

Acq: 27 Feb 2025 14:28

Instrument :

MSVOA_N

ClientSampleId :





PB166880ZHE#01

Tgt Ion: 62 Resp: 1251

Ion Ratio Lower Upper

62 100

98 0.0 0.0 19.6



#43

Isopropyl Acetate

Concen: 32.803 ug/l

RT: 8.688 min Scan# 1145

Delta R.T. 0.005 min

Lab File: VN085888.D

Acq: 27 Feb 2025 14:28




Tgt Ion: 43 Resp: 175654


Ion Ratio Lower Upper

43 100

61 20.9 22.6 34.0#

87 14.3 11.8 17.8





#45

1,2-Dichloropropane

Concen: 0.849 ug/l

RT: 9.747 min Scan# 11

Delta R.T. 0.129 min

Lab File: VN085888.D

Acq: 27 Feb 2025 14:28

Instrument :

MSVOA_N

ClientSampleId :





PB166880ZHE#01

Tgt Ion: 63 Resp: 2421

Ion Ratio Lower Upper

63 100

65 0.0 25.1 37.7#



#48

Methyl methacrylate

Concen: 64.954 ug/l

RT: 9.741 min Scan# 1324

Delta R.T. 0.065 min

Lab File: VN085888.D

Acq: 27 Feb 2025 14:28




Tgt Ion: 41 Resp: 146346


Ion Ratio Lower Upper

41 100

69 0.1 76.0 114.0#

39 47.7 42.3 63.5





#49

1,4-Dioxane

Concen: 1.284 ug/l

RT: 9.735 min Scan# 11

Delta R.T. 0.047 min

Lab File: VN085888.D

Acq: 27 Feb 2025 14:28

Instrument :

MSVOA_N

ClientSampleId :

PB166880ZHE#01





Tgt Ion: 88 Resp: 55

Ion Ratio Lower Upper

88 100

43 2934574.5 18.6 27.8#

58 16930.9 56.7 85.1#



#50

Toluene-d8

Concen: 53.735 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. 0.006 min

Lab File: VN085888.D




Acq: 27 Feb 2025 14:28


Tgt Ion: 98 Resp: 509258

Ion Ratio Lower Upper

98 100

100 64.8 52.1 78.1









#51
4-Methyl-2-Pentanone
Concen: 4.256 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.006 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28




Instrument :
MSVOA_N
ClientSampleId :
PB166880ZHE#01


Tgt Ion: 43 Resp: 12448
Ion Ratio Lower Upper
43 100
58 27.8 34.6 52.0#



#53
t-1,3-Dichloropropene
Concen: 0.345 ug/l
RT: 10.718 min Scan# 1490
Delta R.T. -0.111 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28

Tgt Ion: 75 Resp: 1372
Ion Ratio Lower Upper
75 100
77 62.0 25.4 38.0#









#54
 cis-1,3-Dichloropropene
 Concen: 15.257 ug/l
 RT: 10.359 min Scan# 1429
 Delta R.T. 0.053 min
 Lab File: VN085888.D
 Acq: 27 Feb 2025 14:28




Instrument :
 MSVOA_N
 ClientSampleId :
 PB166880ZHE#01


Tgt Ion: 75 Resp: 66583
 Ion Ratio Lower Upper
 75 100
 77 0.8 26.1 39.1#
 39 24.0 33.4 50.0#



#56
 Ethyl methacrylate
 Concen: 1.159 ug/l
 RT: 10.977 min Scan# 1534
 Delta R.T. 0.106 min
 Lab File: VN085888.D
 Acq: 27 Feb 2025 14:28

Tgt Ion: 69 Resp: 2482
 Ion Ratio Lower Upper
 69 100
 41 3016.9 46.0 69.0#
 39 1208.9 22.0 33.0#









#57
1,3-Dichloropropane
Concen: 2.017 ug/l
RT: 11.200 min Scan# 1565
Delta R.T. 0.041 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28




Instrument :
MSVOA_N
ClientSampleId :
PB166880ZHE#01


Tgt Ion: 76 Resp: 9271
Ion Ratio Lower Upper
76 100
78 0.0 25.9 38.9#



#59
2-Hexanone
Concen: 46.352 ug/l
RT: 11.200 min Scan# 1572
Delta R.T. 0.012 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28

Tgt Ion: 43 Resp: 93271
Ion Ratio Lower Upper
43 100
58 26.4 29.8 89.5#









#62
4-Bromofluorobenzene
Concen: 49.756 ug/l
RT: 12.847 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28




Instrument :
MSVOA_N
ClientSampleId :
PB166880ZHE#01


Tgt Ion: 95 Resp: 155365
Ion Ratio Lower Upper
95 100
174 77.1 0.0 152.4
176 73.2 0.0 146.6



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1685
Delta R.T. 0.006 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28

Tgt Ion: 117 Resp: 371093
Ion Ratio Lower Upper
117 100
82 58.1 45.7 68.5
119 32.5 26.2 39.2









#70
Styrene
Concen: 0.592 ug/l
RT: 12.412 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28




Instrument :
MSVOA_N
ClientSampleId :
PB166880ZHE#01


Tgt Ion:104 Resp: 64
Ion Ratio Lower Upper
104 100
78 84.4 39.0 58.6#
103 0.0 43.3 64.9#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.006 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28

Tgt Ion:152 Resp: 134346
Ion Ratio Lower Upper
152 100
115 60.7 30.4 91.3
150 155.9 0.0 345.4









#74
N-amy1 acetate
Concen: 0.496 ug/l
RT: 12.482 min Scan# 1
Delta R.T. -0.006 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28




Instrument :
MSVOA_N
ClientSampleId :
PB166880ZHE#01


Tgt Ion: 43	Resp: 1611
Ion Ratio	Lower Upper
43	100
70	8.3 41.0 61.4#
55	0.0 24.6 36.8#
61	0.0 21.6 32.4#



#76
1,2,3-Trichloropropane
Concen: 24.310 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. -0.141 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28

Tgt Ion: 75	Resp: 72547
Ion Ratio	Lower Upper
75	100
77	1.3 100.0 300.0#









#81
trans-1,4-Dichloro-2-butene
Concen: 70.930 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.118 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28




Instrument :
MSVOA_N
ClientSampleId :
PB166880ZHE#01


Tgt Ion: 75 Resp: 72547
Ion Ratio Lower Upper
75 100
53 0.0 74.2 111.4#
89 0.0 34.9 52.3#



#84
1,2,4-Trimethylbenzene
Concen: 0.213 ug/l
RT: 13.482 min Scan# 1960
Delta R.T. 0.006 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28

Tgt Ion: 105 Resp: 1579
Ion Ratio Lower Upper
105 100
120 40.3 22.4 67.3









#86
p-Isopropyltoluene
Concen: 0.633 ug/l
RT: 13.953 min Scan# 2001
Delta R.T. 0.230 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28

Instrument :
MSVOA_N
ClientSampleId :
PB166880ZHE#01

Tgt Ion:119 Resp: 59
Ion Ratio Lower Upper
119 100
134 0.0 13.0 39.0#
91 0.0 11.7 35.1#



#95
Naphthalene
Concen: 0.362 ug/l
RT: 15.647 min Scan# 2328
Delta R.T. 0.012 min
Lab File: VN085888.D
Acq: 27 Feb 2025 14:28

Tgt Ion:128 Resp: 2121
Ion Ratio Lower Upper
128 100
127 8.9 10.2 15.2#
129 0.0 8.7 13.1#

