

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022725\
 Data File : VN085881.D
 Acq On : 27 Feb 2025 10:53
 Operator : JC\MD
 Sample : VN0227MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0227MBL01

Quant Time: Feb 28 01:56:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	179265	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	337379	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	305756	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	122242	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	142961	61.512	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	123.020%
35) Dibromofluoromethane	8.165	113	128062	58.008	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	116.020%
50) Toluene-d8	10.565	98	417746	52.010	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.020%
62) 4-Bromofluorobenzene	12.847	95	121363	45.860	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	91.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022725\
Data File : VN085881.D
Acq On : 27 Feb 2025 10:53
Operator : JC\MD
Sample : VN0227MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0227MBL01

Quant Time: Feb 28 01:56:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

