

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022725\
 Data File : VN085886.D
 Acq On : 27 Feb 2025 13:39
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Feb 28 01:59:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	266732	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	489018	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	429308	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	189315	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	171253	49.522	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.040%
35) Dibromofluoromethane	8.165	113	157468	49.210	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.420%
50) Toluene-d8	10.565	98	543422	46.677	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.360%
62) 4-Bromofluorobenzene	12.847	95	191863	50.018	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	100.040%
Target Compounds						
					Qvalue	
78) n-propylbenzene	13.035	91	6213	0.418	ug/l	99
84) 1,2,4-Trimethylbenzene	13.476	105	11603	1.110	ug/l	99
95) Naphthalene	15.641	128	19472	2.358	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022725\
Data File : VN085886.D
Acq On : 27 Feb 2025 13:39
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Feb 28 01:59:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

