

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022725\
 Data File : VN085887.D
 Acq On : 27 Feb 2025 14:04
 Operator : JC\MD
 Sample : PB166880TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166880TB

Quant Time: Feb 28 02:00:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	256358	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	480894	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	460446	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	178126	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	168813	50.792	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 101.580%		
35) Dibromofluoromethane	8.165	113	156118	49.613	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.220%		
50) Toluene-d8	10.565	98	585113	51.107	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.220%		
62) 4-Bromofluorobenzene	12.847	95	190177	50.416	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 100.840%		
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	56286	53.522	ug/l	98
20) Methylene Chloride	5.265	84	4623	1.367	ug/l #	86
37) Ethyl Acetate	7.565	43	4607	1.257	ug/l	96
43) Isopropyl Acetate	8.689	43	177708	27.382	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022725\
Data File : VN085887.D
Acq On : 27 Feb 2025 14:04
Operator : JC\MD
Sample : PB166880TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166880TB

Quant Time: Feb 28 02:00:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

