

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022725\
 Data File : VN085889.D
 Acq On : 27 Feb 2025 14:53
 Operator : JC\MD
 Sample : PB166880ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166880ZHE#02

Quant Time: Feb 28 02:01:13 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	205149	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	387021	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	370328	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	138487	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	147141	55.323	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	110.640%	
35) Dibromofluoromethane	8.165	113	133257	52.619	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	105.240%	
50) Toluene-d8	10.565	98	496451	53.880	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	107.760%	
62) 4-Bromofluorobenzene	12.847	95	151615	49.943	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	99.880%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	44228	52.554	ug/l	95
20) Methylene Chloride	5.265	84	3713	1.372	ug/l #	81
37) Ethyl Acetate	7.571	43	4168	1.413	ug/l #	90
43) Isopropyl Acetate	8.688	43	182625	35.118	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022725\
Data File : VN085889.D
Acq On : 27 Feb 2025 14:53
Operator : JC\MD
Sample : PB166880ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166880ZHE#02

Quant Time: Feb 28 02:01:13 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

