

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022725\
 Data File : VN085893.D
 Acq On : 27 Feb 2025 16:30
 Operator : JC\MD
 Sample : Q1427-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-227

Quant Time: Feb 28 02:03:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	222509	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	438081	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	411277	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	160017	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	156778	54.347	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 108.700%		
35) Dibromofluoromethane	8.165	113	138696	48.384	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 96.760%		
50) Toluene-d8	10.565	98	525192	50.356	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 100.720%		
62) 4-Bromofluorobenzene	12.847	95	174986	50.923	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 101.840%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	62180	68.121	ug/l	93
20) Methylene Chloride	5.271	84	15250	5.195	ug/l	94
43) Isopropyl Acetate	8.688	43	130446	21.957	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022725\
Data File : VN085893.D
Acq On : 27 Feb 2025 16:30
Operator : JC\MD
Sample : Q1427-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-227

Quant Time: Feb 28 02:03:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration

