

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022725\
 Data File : VN085882.D
 Acq On : 27 Feb 2025 11:17
 Operator : JC\MD
 Sample : VN0227WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0227WBL01

Quant Time: Feb 28 01:56:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Feb 19 03:43:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	204202	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	390858	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	343319	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	124825	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	150865	56.986	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.980%
35) Dibromofluoromethane	8.165	113	138357	54.097	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.200%
50) Toluene-d8	10.565	98	445550	47.881	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.760%
62) 4-Bromofluorobenzene	12.847	95	135219	44.104	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	88.200%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN022725\
Data File : VN085882.D
Acq On : 27 Feb 2025 11:17
Operator : JC\MD
Sample : VN0227WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0227WBL01

Quant Time: Feb 28 01:56:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N021825W.M
Quant Title : SW846 8260
QLast Update : Wed Feb 19 03:43:32 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085882.D
Acq: 27 Feb 2025 11:17




Instrument :
MSVOA_N
ClientSampleId :
VN0227WBL01


Tgt Ion:168 Resp: 204202
Ion Ratio Lower Upper
168 100
99 59.2 47.9 71.9



#33
1,2-Dichloroethane-d4
Concen: 56.986 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. 0.006 min
Lab File: VN085882.D
Acq: 27 Feb 2025 11:17

Tgt Ion: 65 Resp: 150865
Ion Ratio Lower Upper
65 100
67 54.0 0.0 106.2





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1

Delta R.T. 0.006 min

Lab File: VN085882.D

Acq: 27 Feb 2025 11:17

Instrument :

MSVOA_N

ClientSampleId :

VN0227WBL01





Tgt Ion:114 Resp: 390858

Ion Ratio Lower Upper

114 100

63 20.3 0.0 39.2

88 16.2 0.0 30.0



#35

Dibromofluoromethane

Concen: 54.097 ug/l

RT: 8.165 min Scan# 1056

Delta R.T. 0.006 min

Lab File: VN085882.D

Acq: 27 Feb 2025 11:17




Tgt Ion:113 Resp: 138357


Ion Ratio Lower Upper

113 100

111 101.5 82.2 123.4

192 19.1 14.7 22.1









#50
Toluene-d8
Concen: 47.881 ug/l
RT: 10.565 min Scan# 1464
Delta R.T. 0.006 min
Lab File: VN085882.D
Acq: 27 Feb 2025 11:17




Instrument :
MSVOA_N
ClientSampleId :
VN0227WBL01


Tgt Ion: 98 Resp: 445550
Ion Ratio Lower Upper
98 100
100 63.4 52.1 78.1



#62
4-Bromofluorobenzene
Concen: 44.104 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.006 min
Lab File: VN085882.D
Acq: 27 Feb 2025 11:17

Tgt Ion: 95 Resp: 135219
Ion Ratio Lower Upper
95 100
174 79.3 0.0 152.4
176 74.9 0.0 146.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 117
Delta R.T. 0.006 min
Lab File: VN085882.D
Acq: 27 Feb 2025 11:17

Instrument :
MSVOA_N
ClientSampleId :
VN0227WBL01

Tgt Ion:117 Resp: 343319
Ion Ratio Lower Upper
117 100
82 54.8 45.7 68.5
119 32.5 26.2 39.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.788 min Scan# 2012
Delta R.T. -0.000 min
Lab File: VN085882.D
Acq: 27 Feb 2025 11:17

Tgt Ion:152 Resp: 124825
Ion Ratio Lower Upper
152 100
115 59.7 30.4 91.3
150 158.9 0.0 345.4

