

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022819\
 Data File : VN054012.D
 Acq On : 28 Feb 2019 18:47
 Operator : JC/SP
 Sample : K1647-05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Mar 01 03:58:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 28 04:05:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	592624	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1093752	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1074224	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	401521	50.00	ug/l	0.00


System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.03	65	383637	55.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.44%	
35) Dibromofluoromethane	7.59	113	352501	49.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.68%	
50) Toluene-d8	10.09	98	1389713	52.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.56%	
62) 4-Bromofluorobenzene	12.40	95	465363	53.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.10%	

Target Compounds

					Qvalue
8) Diethyl Ether	3.41	74	955	0.263	ug/l 59
11) Tert butyl alcohol	4.78	59	341	0.774	ug/l # 73
14) Allyl chloride	4.56	41	12218	1.191	ug/l # 27
16) Acetone	3.82	43	18350	Below Cal	89
18) Methyl Acetate	4.34	43	3410	0.702	ug/l # 62
20) Methylene Chloride	4.55	84	300110	43.649	ug/l 89
25) 2-Butanone	6.84	43	1948	0.714	ug/l # 84
29) Tetrahydrofuran	7.22	42	1570	0.864	ug/l # 57
31) Cyclohexane	7.77	56	368	Below Cal	95
36) 1,1-Dichloropropene	7.66	75	41257	4.056	ug/l # 47
38) Carbon Tetrachloride	7.67	117	50495	4.773	ug/l # 18
43) Isopropyl Acetate	8.17	43	33437	2.727	ug/l # 88
45) 1,2-Dichloropropane	9.27	63	1932	0.229	ug/l # 44
48) Methyl methacrylate	9.18	41	8972	1.591	ug/l # 12
49) 1,4-Dioxane	9.27	88	104	1.361	ug/l # 1
51) 4-Methyl-2-Pentanone	10.02	43	8824	1.400	ug/l # 34
54) cis-1,3-Dichloropropene	9.90	75	63368	5.356	ug/l # 69
56) Ethyl methacrylate	10.54	69	1061	2.068	ug/l # 1
57) 1,3-Dichloropropane	10.76	76	12631	1.088	ug/l # 43
59) 2-Hexanone	10.76	43	181904	44.385	ug/l # 48
71) Bromoform	12.40	173	3672	0.646	ug/l # 1
74) N-amyl acetate	11.88	43	18443	2.251	ug/l # 50
76) 1,2,3-Trichloropropane	12.40	75	227147	35.265	ug/l # 100
80) 1,3,5-Trimethylbenzene	12.73	105	7348	0.287	ug/l 91
81) trans-1,4-Dichloro-2-buten	12.40	75	227147	111.460	ug/l # 15
84) 1,2,4-Trimethylbenzene	13.04	105	9876	0.387	ug/l 93
95) Naphthalene	15.14	128	4038	3.449	ug/l # 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN054012.D

Acq: 28 Feb 2019 18:47


Instrument :
MSVOA_N
ClientSampled :

Tot Ion:168 Resp: 592624

Ion Ratio Lower Upper

168 100

99 52.1 36.9 55.3



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.67

250000

200000



150000

100000

50000

0

Time--> 7.50 7.60 7.70 7.80 7.90



#8

Diethyl Ether

Concen: 0.263 ug/l

RT: 3.41 min Scan# 510

Delta R.T. 0.00 min

Lab File: VN054012.D


Acq: 28 Feb 2019 18:47

Tgt Ion: 74 Resp: 955

Ion Ratio Lower Upper

74 100

45 133.2 46.9 140.6



Abundance Ion 74.00 (73.70 to 74.70): VN

Ion 45.00 (44.70 to 45.70): VN

800


600


400

200

0

Time--> 3.40 3.45









#11

Tert butyl alcohol
 Concen: 0.774 ug/l
 RT: 4.78 min Scan# 935
 Delta R.T. 0.00 min
 Lab File: VN054012.D
 Acq: 28 Feb 2019 18:47

Instrument :
 MSVOA_N
 ClientSampleId :




Tot Ion: 59 Resp: 341
 Ion Ratio Lower Upper
 59 100
 57 0.0 8.1 12.1#




#14

Allyl chloride
 Concen: 1.191 ug/l
 RT: 4.56 min Scan# 866
 Delta R.T. 0.23 min
 Lab File: VN054012.D
 Acq: 28 Feb 2019 18:47

Tgt Ion: 41 Resp: 12218
 Ion Ratio Lower Upper
 41 100
 39 0.8 48.8 73.2#
 76 0.0 30.9 46.3#





#16

Acetone

Concen: Below Cal

RT: 3.82 min Scan# 638

Delta R.T. 0.01 min

Lab File: VN054012.D

Acq: 28 Feb 2019 18:47





Instrument :
MSVOA_N
ClientSampled :

Tot Ion: 43 Resp: 18350

Ion Ratio Lower Upper

43 100

58 28.0 27.6 41.4



#18

Methyl Acetate

Concen: 0.702 ug/l

RT: 4.34 min Scan# 797

Delta R.T. 0.01 min

Lab File: VN054012.D




Acq: 28 Feb 2019 18:47


Tgt Ion: 43 Resp: 3410

Ion Ratio Lower Upper

43 100

74 6.2 20.5 30.7#









#20
Methylene Chloride
Concen: 43.649 ug/l
RT: 4.55 min Scan# 865
Delta R.T. 0.00 min
Lab File: VN054012.D
Acq: 28 Feb 2019 18:47




Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 84 Resp: 300110
Ion Ratio Lower Upper
84 100
49 128.9 89.7 134.5
51 40.3 27.8 41.8
86 62.5 51.3 76.9



#25
2-Butanone
Concen: 0.714 ug/l
RT: 6.84 min Scan# 1577
Delta R.T. 0.01 min
Lab File: VN054012.D
Acq: 28 Feb 2019 18:47

Tgt Ion: 43 Resp: 1948
Ion Ratio Lower Upper
43 100
72 19.4 22.1 33.1#





#29

Tetrahydrofuran

Concen: 0.864 ug/l

RT: 7.22 min Scan# 1694

Delta R.T. 0.01 min

Lab File: VN054012.D

Acq: 28 Feb 2019 18:47

Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 42 Resp: 1570

Ion Ratio Lower Upper




42 100

72 12.5 36.9 55.3#

71 20.8 34.6 52.0#



Abundance Ion 42.00 (41.70 to 42.70): VN0
Ion 72.00 (71.70 to 72.70): VN0
Ion 71.00 (70.70 to 71.70): VN0



#31

Cyclohexane

Concen: Below Cal

RT: 7.77 min Scan# 1865

Delta R.T. 0.12 min

Lab File: VN054012.D

Acq: 28 Feb 2019 18:47


Tgt Ion: 56 Resp: 368

Ion Ratio Lower Upper



56 100


69 36.7 26.0 39.0

84 90.3 69.1 103.7



Abundance Ion 56.00 (55.70 to 56.70): VN0
Ion 69.00 (68.70 to 69.70): VN0
Ion 84.00 (83.70 to 84.70): VN0









#33
 1,2-Dichloroethane-d4
 Concen: 55.224 ug/l
 RT: 8.03 min Scan# 1945
 Delta R.T. 0.00 min
 Lab File: VN054012.D
 Acq: 28 Feb 2019 18:47




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion: 65 Resp: 383637
 Ion Ratio Lower Upper
 65 100
 67 54.6 0.0 111.8



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN054012.D
 Acq: 28 Feb 2019 18:47

Tgt Ion: 114 Resp: 1093752
 Ion Ratio Lower Upper
 114 100
 63 21.0 0.0 37.4
 88 16.0 0.0 30.4









#35
 Dibromofluoromethane
 Concen: 49.839 ug/l
 RT: 7.59 min Scan# 1809
 Delta R.T. 0.00 min
 Lab File: VN054012.D
 Acq: 28 Feb 2019 18:47




Instrument :
 MSVOA_N
 ClientSampleId :


Tot Ion:113 Resp: 352501
 Ion Ratio Lower Upper
 113 100
 111 102.3 81.1 121.7
 192 20.8 18.8 28.2



#36
 1,1-Dichloropropene
 Concen: 4.056 ug/l
 RT: 7.66 min Scan# 1832
 Delta R.T. -0.13 min
 Lab File: VN054012.D
 Acq: 28 Feb 2019 18:47

Tgt Ion: 75 Resp: 41257
 Ion Ratio Lower Upper
 75 100
 110 7.9 19.0 57.0#
 77 0.0 25.4 38.0#





#38

Carbon Tetrachloride

Concen: 4.773 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. -0.11 min

Lab File: VN054012.D

Acq: 28 Feb 2019 18:47

Instrument :
MSVOA_N
ClientSampleId :


Tot Ion:117 Resp: 50495

Ion Ratio Lower Upper




117 100

119 7.0 76.2 114.2#

121 0.0 24.6 36.8#



Abundance Ion 116.80 (116.50 to 117.50): V
Ion 118.80 (118.50 to 119.50): V
Ion 120.80 (120.50 to 121.50): V



#43

Isopropyl Acetate

Concen: 2.727 ug/l

RT: 8.17 min Scan# 1990

Delta R.T. 0.01 min

Lab File: VN054012.D

Acq: 28 Feb 2019 18:47


Tgt Ion: 43 Resp: 33437

Ion Ratio Lower Upper



43 100


61 16.1 17.4 26.0#

87 9.5 11.5 17.3#



Abundance Ion 43.00 (42.70 to 43.70): V
Ion 61.00 (60.70 to 61.70): V
Ion 87.00 (86.70 to 87.70): V





#45

1,2-Dichloropropane

Concen: 0.229 ug/l

RT: 9.27 min Scan# 2331

Delta R.T. 0.15 min

Lab File: VN054012.D

Acq: 28 Feb 2019 18:47

Instrument :

MSVOA_N


ClientSampleId :

Tot Ion: 63 Resp: 1932

Ion Ratio Lower Upper




63 100

65 0.0 24.3 36.5#



Abundance Ion 62.90 (62.60 to 63.60): VNC

Ion 64.90 (64.60 to 65.60): VNC



#48

Methyl methacrylate

Concen: 1.591 ug/l

RT: 9.18 min Scan# 2305

Delta R.T. -0.02 min

Lab File: VN054012.D

Acq: 28 Feb 2019 18:47


Tgt Ion: 41 Resp: 8972

Ion Ratio Lower Upper

41 100

69 0.0 74.8 112.2#



39 0.0 39.0 58.6#




Abundance Ion 41.00 (40.70 to 41.70): VNC

Ion 69.00 (68.70 to 69.70): VNC

Ion 39.00 (38.70 to 39.70): VNC





#49

1,4-Dioxane

Concen: 1.361 ug/l

RT: 9.27 min Scan# 2332

Delta R.T. 0.07 min

Lab File: VN054012.D

Acq: 28 Feb 2019 18:47

Instrument :
MSVOA_N
ClientSampled :


Tot Ion: 88 Resp: 104

Ion Ratio Lower Upper




88 100

43 1548144.2 21.2 31.8#

58 8457.7 54.2 81.2#



Abundance Ion 88.00 (87.70 to 88.70): VN0
Ion 43.00 (42.70 to 43.70): VN0
Ion 58.00 (57.70 to 58.70): VN0



#50

Toluene-d8

Concen: 52.782 ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN054012.D


Acq: 28 Feb 2019 18:47

Tgt Ion: 98 Resp: 1389713



Ion Ratio Lower Upper


98 100

100 63.9 51.9 77.9



Abundance Ion 98.00 (97.70 to 98.70): VN0
Ion 100.00 (99.70 to 100.70): VN0





#51

4-Methyl-2-Pentanone

Concen: 1.400 ug/l

RT: 10.02 min Scan# 2565

Delta R.T. 0.04 min

Lab File: VN054012.D

Acq: 28 Feb 2019 18:47

Instrument :

MSVOA_N





ClientSampled :

Tot Ion: 43 Resp: 8824

Ion Ratio Lower Upper

43 100

58 0.0 32.7 49.1#



#54

cis-1,3-Dichloropropene

Concen: 5.356 ug/l

RT: 9.90 min Scan# 2529

Delta R.T. 0.06 min

Lab File: VN054012.D

Acq: 28 Feb 2019 18:47




Tgt Ion: 75 Resp: 63368


Ion Ratio Lower Upper

75 100

77 0.6 25.8 38.6#

39 33.1 31.4 47.0





#56

Ethyl methacrylate

Concen: 2.068 ug/l

RT: 10.54 min Scan# 2728

Delta R.T. 0.12 min

Lab File: VN054012.D

Acq: 28 Feb 2019 18:47





Instrument :

MSVOA_N

ClientSampled :

Tot Ion: 69 Resp: 1061

Ion	Ratio	Lower	Upper
69	100		
41	4642.6	47.5	71.3#
39	1915.3	23.4	35.0#



#57

1,3-Dichloropropane

Concen: 1.088 ug/l

RT: 10.76 min Scan# 2796




Delta R.T. 0.06 min


Lab File: VN054012.D

Acq: 28 Feb 2019 18:47

Tgt Ion: 76 Resp: 12631

Ion	Ratio	Lower	Upper
76	100		
78	0.0	25.2	37.8#





#59

2-Hexanone

Concen: 44.385 ug/l

RT: 10.76 min Scan# 2796

Delta R.T. 0.01 min

Lab File: VN054012.D

Acq: 28 Feb 2019 18:47

Instrument :

MSVOA_N





ClientSampled :

Tot Ion: 43 Resp: 181904

Ion Ratio Lower Upper

43 100

58 19.8 29.1 87.4#



#62

4-Bromofluorobenzene

Concen: 53.051 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.00 min

Lab File: VN054012.D

Acq: 28 Feb 2019 18:47




Tgt Ion: 95 Resp: 465363


Ion Ratio Lower Upper

95 100

174 86.8 0.0 188.8

176 84.0 0.0 182.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2997
Delta R.T. 0.00 min
Lab File: VN054012.D
Acq: 28 Feb 2019 18:47

Instrument :
MSVOA_N
ClientSampled :

Tot Ion:117 Resp: 1074224
Ion Ratio Lower Upper
117 100
82 54.4 41.4 62.2
119 32.0 25.3 37.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V






#71
Bromoform
Concen: 0.646 ug/l
RT: 12.40 min Scan# 3305
Delta R.T. 0.27 min
Lab File: VN054012.D
Acq: 28 Feb 2019 18:47

Tgt Ion:173 Resp: 3672
Ion Ratio Lower Upper
173 100
175 678.0 24.3 72.8#
254 0.0 0.2 0.2#



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3599

Delta R.T. 0.00 min

Lab File: VN054012.D

Acq: 28 Feb 2019 18:47

Instrument :

MSVOA_N

ClientSampled :


Tot Ion:152 Resp: 401521

Ion Ratio Lower Upper

152 100

115 57.7 27.2 81.6




150 156.4 0.0 348.2



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 2.251 ug/l

RT: 11.88 min Scan# 3142

Delta R.T. -0.19 min

Lab File: VN054012.D

Acq: 28 Feb 2019 18:47

Tgt Ion: 43 Resp: 18443


Ion Ratio Lower Upper

43 100

70 2.3 37.4 56.2#

55 6.3 21.4 32.0#

61 9.8 20.9 31.3#






Abundance Ion 43.00 (42.70 to 43.70): V

Ion 70.00 (69.70 to 70.70): V

Ion 55.00 (54.70 to 55.70): V

Ion 61.00 (60.70 to 61.70): V









#76
 1,2,3-Trichloropropane
 Concen: 35.265 ug/l
 RT: 12.40 min Scan# 3306
 Delta R.T. -0.15 min
 Lab File: VN054012.D
 Acq: 28 Feb 2019 18:47




Instrument :
 MSVOA_N
 ClientSampled :


Tot Ion: 75 Resp: 227147
 Ion Ratio Lower Upper
 75 100
 77 1.3 0.0 0.0#



#80
 1,3,5-Trimethylbenzene
 Concen: 0.287 ug/l
 RT: 12.73 min Scan# 3409
 Delta R.T. 0.00 min
 Lab File: VN054012.D
 Acq: 28 Feb 2019 18:47

Tgt Ion: 105 Resp: 7348
 Ion Ratio Lower Upper
 105 100
 120 56.5 25.1 75.2





#81

trans-1,4-Dichloro-2-butene

Concen: 111.460 ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.10 min

Lab File: VN054012.D

Acq: 28 Feb 2019 18:47

Instrument :

MSVOA_N

ClientSampled :


Tot Ion: 75 Resp: 227147

Ion Ratio Lower Upper

75 100

53 0.1 68.6 102.8#




89 0.0 35.3 52.9#



Abundance Ion 74.90 (74.60 to 75.60): VN0

Ion 53.00 (52.70 to 53.70): VN0

Ion 88.90 (88.60 to 89.60): VN0



#84

1,2,4-Trimethylbenzene

Concen: 0.387 ug/l

RT: 13.04 min Scan# 3504

Delta R.T. 0.00 min

Lab File: VN054012.D


Acq: 28 Feb 2019 18:47

Tgt Ion:105 Resp: 9876

Ion Ratio Lower Upper



105 100


120 51.1 23.2 69.5



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V





#95

Naphthalene

Concen: 3.449 ug/l

RT: 15.14 min Scan# 4156

Delta R.T. 0.00 min

Lab File: VN054012.D

Acq: 28 Feb 2019 18:47

Instrument :

MSVOA_N

ClientSampleId :


Tot Ion:128 Resp: 4038

Ion Ratio Lower Upper

128 100

127 5.5 10.2 15.2#

129 9.5 8.6 12.8



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

