

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022819\
 Data File : VN053997.D
 Acq On : 28 Feb 2019 12:06
 Operator : JC/SP
 Sample : PB117454TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB117454TB

Quant Time: Mar 01 04:19:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 28 04:05:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	587375	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1079251	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1053753	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	350041	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.02	65	380466	55.26	ug/l	0.00
Spiked Amount						
			Recovery	=	110.52%	
35) Dibromofluoromethane	7.59	113	346849	49.70	ug/l	0.00
Spiked Amount						
			Recovery	=	99.40%	
50) Toluene-d8	10.09	98	1393863	53.65	ug/l	0.00
Spiked Amount						
			Recovery	=	107.30%	
62) 4-Bromofluorobenzene	12.40	95	440995	50.95	ug/l	0.00
Spiked Amount						
			Recovery	=	101.90%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022819\
Data File : VN053997.D
Acq On : 28 Feb 2019 12:06
Operator : JC/SP
Sample : PB117454TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB117454TB

Quant Time: Mar 01 04:19:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 28 04:05:45 2019
Response via : Initial Calibration

