

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022819\
 Data File : VN054004.D
 Acq On : 28 Feb 2019 15:13
 Operator : JC/SP
 Sample : K1349-16
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-03-022719-B

Quant Time: Mar 01 04:47:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 28 04:05:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	580588	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1102466	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1075577	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	404748	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	392521	57.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.34%	
35) Dibromofluoromethane	7.59	113	355178	49.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.64%	
50) Toluene-d8	10.09	98	1398055	52.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.36%	
62) 4-Bromofluorobenzene	12.40	95	461710	52.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.44%	
Target Compounds						
20) Methylene Chloride	4.55	84	616329	91.499	ug/l	89
43) Isopropyl Acetate	8.17	43	38804	3.139	ug/l #	88
84) 1,2,4-Trimethylbenzene	13.04	105	30207	1.173	ug/l	98
95) Naphthalene	15.13	128	13509	4.001	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022819\
Data File : VN054004.D
Acq On : 28 Feb 2019 15:13
Operator : JC/SP
Sample : K1349-16
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-03-022719-B

Quant Time: Mar 01 04:47:34 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 28 04:05:45 2019
Response via : Initial Calibration

