

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022819\
 Data File : VN054005.D
 Acq On : 28 Feb 2019 15:40
 Operator : JC/SP
 Sample : K1349-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampled :
 OR-03-022719-A

Quant Time: Mar 01 04:50:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 28 04:05:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	588878	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1092442	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1053807	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	388225	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	8.03	65	396143	57.39	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	114.78%	
35) Dibromofluoromethane	7.59	113	352163	49.85	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	99.70%	
50) Toluene-d8	10.09	98	1394934	53.04	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	106.08%	
62) 4-Bromofluorobenzene	12.40	95	450497	51.42	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	102.84%	

Target Compounds

					Qvalue	
16) Acetone	3.81	43	27921	3.281	ug/l	94
20) Methylene Chloride	4.55	84	25813	3.778	ug/l	90
43) Isopropyl Acetate	8.17	43	36566	2.985	ug/l #	91
95) Naphthalene	15.14	128	10166	3.831	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022819\
 Data File : VN054005.D
 Acq On : 28 Feb 2019 15:40
 Operator : JC/SP
 Sample : K1349-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-03-022719-A

Quant Time: Mar 01 04:50:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 28 04:05:45 2019
 Response via : Initial Calibration

