

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN022819\
 Data File : VN054009.D
 Acq On : 28 Feb 2019 17:27
 Operator : JC/SP
 Sample : K1651-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 351-281

Quant Time: Mar 01 05:01:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
 Quant Title : SW846 8260
 QLast Update : Thu Feb 28 04:05:45 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	582312	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1083424	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1081850	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	430219	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	385878	56.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.06%	
35) Dibromofluoromethane	7.59	113	346058	49.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.78%	
50) Toluene-d8	10.09	98	1389929	53.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.58%	
62) 4-Bromofluorobenzene	12.40	95	485397	55.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.72%	
Target Compounds						
20) Methylene Chloride	4.55	84	583627	86.388	ug/l	91
43) Isopropyl Acetate	8.18	43	43664	3.595	ug/l #	82
95) Naphthalene	15.13	128	104547	8.955	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN022819\
Data File : VN054009.D
Acq On : 28 Feb 2019 17:27
Operator : JC/SP
Sample : K1651-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
351-281

Quant Time: Mar 01 05:01:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N022519W.M
Quant Title : SW846 8260
QLast Update : Thu Feb 28 04:05:45 2019
Response via : Initial Calibration

